

PALOS VERDES LANDFILL
REMEDIAL INVESTIGATION REPORT

APPENDIX E.10

EMISSIONS TESTS; PALOS VERDES ENERGY
RECOVERY FROM LANDFILL GAS (CARNOT)

**EMISSIONS TESTS ON THE PALOS VERDES
ENERGY RECOVERY FROM LANDFILL GAS
(PVERG) FACILITY, UNIT 1, DECEMBER 1992**

Prepared For:

COUNTY SANITATION DISTRICTS OF LOS ANGELES COUNTY
Whittier, California

For Submittal To:

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT

Prepared By:


Craig H. Fry

CARNOT
Tustin, California

JANUARY 1993

REVIEW AND CERTIFICATION


All work, calculations, and other activities and tasks performed and documented in this report were carried out under my direction and supervision.



Date 2/2/93

Craig H. Fry
Assistant Manager
Emission Measurement Division

I have reviewed, technically and editorially, details, calculations, results, conclusions and other appropriate written material contained herein, and hereby certify that the presented material is authentic and accurate.



Date 2-3-93

Edward J. Filadelfia
Senior Engineer

TABLE OF CONTENTS

<u>SECTION</u>		<u>PAGE</u>
1.0	INTRODUCTION	1
2.0	TEST UNIT DESCRIPTION	4
3.0	TEST DESCRIPTION	5
3.1	TEST CONDITIONS	5
3.2	SAMPLE LOCATIONS	5
3.3	TEST PROCEDURES	7
3.3.1	Methane and Total Gaseous Non-Methane Organics	8
3.3.2	Oxygen, Nitrogen, Carbon Monoxide, Carbon Dioxide and Nitrogen Oxides	8
3.3.3	Moisture	9
3.3.4	Flow Rate	9
3.3.5	Particulate Matter	9
3.3.6	Trace Organic Hydrocarbons	10
3.3.7	Hydrogen Sulfide and Reduced Sulfur Compounds	10
4.0	RESULTS	11
APPENDICES		
A	MEASUREMENT PROCEDURES	A-1
B	QUALITY ASSURANCE	B-1
B.1	Quality Assurance Program Summary	
B.2	ARB Certification	
B.3	Calibration Data	
B.4	Chain of Custody	
C	FIELD AND LAB DATA SUMMARIES	C-1
C.1	Sample Locations	
C.2	CEM Data and Strip Charts	
C.3	Particulate and Calculated Exhaust Flow Rate Data	
C.4	Inlet Moisture and Calculated Flow Rate Data	
C.5	Total Hydrocarbon Lab Results	
C.6	Trace Hydrocarbon and Trace Sulfur Results	
C.7	Unit Operation Data	
D	CALCULATIONS	D-1
D.1	General Emissions Calculations	
D.2	Sample Calculations	

SECTION 1.0

INTRODUCTION

Carnot was contracted by the County Sanitation Districts of Los Angeles County (LACSD) to perform a series of air emission related tests on the Palos Verdes Energy Recovery from Landfill Gas (PVERG) Facility, Unit 1. The tests were required by the South Coast Air Quality Management District (SCAQMD) to satisfy the requirements given in the SCAQMD Rule 1150.2, and the SCAQMD Permit to Construct Application No. 134321. Also, these tests were conducted to satisfy the California EPA (Cal-EPA) Toxic Substance Control requirements for the provisions on ambient air and landfill gas work plan for the Palos Verdes Landfill.

Emission tests were conducted on Unit 1 at the boiler exhaust and the landfill gas fuel supply (inlet). Table 1-1 presents a test matrix of the parameters measured at each location. The emission tests were conducted on December 30, 1992. Testing was performed by Craig Fry, Dave Evans and Robert Madrigal of Carnot. The testing was coordinated by Dr. Moon S. Chung of the LACSD. The boiler was maintained at steady load throughout the testing period by LACSD personnel.

A summary of the emission test results and the applicable permit limits for the parameters measured are presented in Table 1-2. Detailed test results are presented in Section 4.0. A discussion of the results and weather conditions occurring during testing are discussed in Section 5.0. All raw data and calculations can be found in the Appendices.

**TABLE 1-1
TEST MATRIX**

Parameter	Inlet	Exhaust
Oxygen (O ₂)	X	X
Carbon Dioxide (CO ₂)	X	X
Carbon Monoxide (CO)		X
Nitrogen Oxides (NO _x)		X
Nitrogen (N ₂)	X	X
Moisture (H ₂ O)	X	X
Methane (CH ₄)	X	X
Total Gaseous Non-Methane Organics (TGNMO)	X	X
Flow Rate	X	X
Particulate Matter (PM)		X
Trace Organic Compounds	X	X
Reduced Sulfur Compounds	X	
Hydrogen Sulfide (H ₂ S)	X	

**TABLE 1-2
SUMMARY OF TEST RESULTS
PVERG BOILER
DECEMBER 30, 1992**

Parameter	Inlet	Exhaust	Permit Limit	Applicable SCAQMD Rule
O ₂ , % dry	11.2	5.33		
CO ₂ , % dry	16.0	12.92		
N ₂ , % dry	54.5	81.8		
H ₂ O, %	4.0	14.5		
Flow Rate, scfm	3,722			
Flow Rate, dscfm	3,573	14,615		
Temperature, °F		337		
NO _x :				
ppm		18.4		
ppm @ 3% O ₂		21.1	60	P/C
lb/hr (as NO ₂)		1.95		
lb/day (as NO ₂)		*	149	P/C
CO:				
ppm		13.90	2,000	407
ppm @ 3% O ₂		16.0		
lb/hr		0.90		
lb/day		*	152	P/C
HC:				
CH ₄ ppm	195,500	ND < 1		
TGNMO ppm	7,200	16.1		
TGNMO lb/hr (as CH ₄)	65.07	0.593		
TGNMO lb/day (as CH ₄)		*	65	P/C
Destruction Eff. %		99.09		
Particulate:				
Total Particulate gr/dscf		0.0027	0.01	475
gr/dscf @ 3% O ₂		0.0056		
gr/dscf @ 12% CO ₂		0.0025		
lb/hr		0.34		
lb/day		*	88	P/C
Total Sulfur, ppm	< 26.2		800 Inlet	431.1

NOTE: The results in this table are the averages of all measurements. See Section 4.0 for complete emission test results.

* - not measured.

P/C - SCAQMD Permit Operating Condition Application No. 134321.

SECTION 2.0

TEST UNIT DESCRIPTION

The Palos Verdes Energy Recovery from Landfill Gas Facility (PVERG) consists of two identical Zurn Keystone landfill gas fired boilers. Unit 1 was tested in this emission test series. The steam production of both boilers is used to drive a steam turbine generator with a maximum electrical output of 13.0 megawatts. Each boiler fires approximately 3,600 scfm of landfill gas with supplemental natural gas which makes up 25% of the total Btu content. The natural gas is used because of the low Btu value (200 Btu/scf) of the inactive landfill gas fuel.

SECTION 3.0

TEST DESCRIPTION

3.1 TEST CONDITIONS

LACSD personnel maintained steady load operations throughout the test period. The boiler was set to run at nominal load. Temperature, fuel flow rate and fuel Btu values were monitored by LACSD throughout the test periods. Table 3-1 summarizes the boiler operating conditions.

**TABLE 3-1
PVERG TEST OPERATING CONDITIONS**

Parameter	Condition
Boiler Excess Oxygen	5.3%
Landfill Gas Flow Rate	3,720 scfm
Natural Gas Flow Rate	209 scfm
Landfill Gas Heating Value	204 Btu/scf
Generator Output	9.2 MW

3.2 SAMPLE LOCATIONS

Samples were collected from two locations in the landfill gas/boiler system. The locations were 1) the landfill fuel gas supplies and 2) the boiler exhausts. Access for sampling in the landfill gas fuel supply (inlet) was gained by removing a pipe plug downstream of the blower outlet and upstream of the boiler inlet. No isokinetic sampling was required at this location.

Samples from the boiler exhaust were collected via two ports in the stack wall. The ports are 4 inch I.D. and are located 90° apart on the circumference of the stack. The ports are 22 feet above the nearest flow disturbance and 5 feet, 5 inches below the stack top. Figure 3-1 is a schematic drawing of the PVERG stack sampling location. All tests requiring isokinetic sampling were conducted using 16




Figure 3-1. PVERG Unit 1, Stack

point traverses of the sample plane.

3.3 TEST PROCEDURES

The test procedures followed for the landfill gas and boiler exhaust measurements are summarized below in Tables 3-2 and 3-3, respectively. The selected procedures are consistent with those recommended by SCAQMD in their "Guidelines for Implementation of Rule 1150.1." Brief discussions of each procedure are given below in Sections 3.3.1 - 3.3.7.

**TABLE 3-2
PVERG INLET TEST PROCEDURES**

Parameter	Sample Medium	Analytical Technique	Reference Method
Methane and Total Gaseous Non-Methane Organics	Tedlar Bag	GC/FID	SCAQMD 25.2
O ₂ , CO ₂ and N ₂	Tedlar Bag	GC/TCD	SCAQMD 10.1
Moisture	Impinger Train	Gravimetric	SCAQMD 4.1
Flow Rate		Plant Flow Meter	
Trace Organics	Tedlar Bag	GC/PID/ELCD	EPA T03
H ₂ S	Tedlar Bag	GC/FPD/ELCD	EPA 15
Reduced Sulfur Compounds	Tedlar Bag	GC/FPD/ELCD	EPA 15

**TABLE 3-3
PVERG EXHAUST TEST PROCEDURES**

Parameter	Sample Medium	Analytical Technique	Reference Method
Methane and Total Gaseous Non-Methane Organics	Tedlar Bag	GC/FID	SCAQMD 25.2
O ₂	CEM	Micro Fuel Cell	SCAQMD 100.1
CO ₂	CEM	NDIR	SCAQMD 100.1
NO _x	CEM	Chemiluminescence	SCAQMD 100.1
CO	CEM	NDIR/GFC	SCAQMD 100.1
Moisture	Impinger train	Gravimetric	SCAQMD 4.1
Flow Rate	Pitot Tube	Differential Pressure	SCAQMD 2.1
Particulate Matter	Wet Impingement Train	Gravimetric	SCAQMD 5.1
Trace Organics	Tedlar Bag	GC/PID/ELCD	EPA T03

3.3.1 Methane and Total Gaseous Non-Methane Organics

Methane and total gaseous non-methane organics were measured using the draft SCAQMD Method 25.2. The inlet and exhaust samples were collected with Tedlar bags and analyzed by the TCA/FID. Atmosphere Assessment Associates performed the analyses on these samples within 72 hours after sampling. Inlet and exhaust were sampled simultaneously for a twenty-minute period.

3.3.2 Oxygen, Nitrogen, Carbon Monoxide, Carbon Dioxide and Nitrogen Oxides

Inlet O₂, CO₂ and N₂ measurements were acquired from Tedlar bag samples and analyzed by GC/TCD using SCAQMD Method 10.1.

Measurements of NO_x, CO, O₂ and CO₂ at the exhaust were conducted using SCAQMD Method 100.1 sampling with a continuous emission monitoring system (CEM).

These CEM measurements were made using Carnot's continuous emissions monitoring system described in Appendix A. The system used a stainless steel probe connected to a 15' heater teflon line.

Then the gaseous sample traveled through a moisture knockout cooled with ice and water. A peristaltic pump continuously drained the knock out. The sample then traveled to the ground via teflon line to an additional conditioning and filtering system.

Leak checks were conducted prior to and at the completion of the test project. The leak checks were conducted by operating the sample pump, plugging the probe inlet and all pressure side system exits except for one analyzer rotameter, then measuring the leakage rate on that rotameter.

EPA Protocol 1 Calibration Gas were used for CO and NO_x analyzer calibration. All other gases were manufacturer certified to be $\pm 1\%$ and traceable to NIST. In accordance with SCAQMD Method 100.1 procedures, a pre- and post-test system bias check was conducted for each test run. The system bias check was conducted by delivering zero and span gas to the CEM probe tip and recording the as-found species concentration. No analyzer adjustments were made between these pre- and post-system bias checks. Calculations for the correction of measured system bias and instrument drift were then applied to each test run. The allowable limit of system bias deviation is 5% of instrument range.

Duplicate emissions measurements were performed to determine nitrogen oxides and carbon monoxide emissions, as well as oxygen and carbon dioxide concentrations. The average concentrations were determined during each test for a period of 64 minutes. This test average was then corrected for measured system bias and drift. Exhaust N₂ data was calculated by difference from the concentration of the other major exhaust gas components.

3.3.3 Moisture

Moisture was measured in both the inlet and outlet streams of the boiler. SCAQMD Method 4.1 was used for the determinations.

Duplicate inlet measurements were made with a separate Method 4.1 train.

The moisture determinations of the flare exhaust were made in combination with the particulate matter tests.

3.3.4 Flow Rate

Landfill gas flow rate into the boiler was set to specification using existing LACSD instrumentation on site, and was monitored and recorded throughout the test period. The inlet gas density is calculated using the concentrations of oxygen, carbon dioxide and methane. Outlet gas flow rate was determined in conjunction with all isokinetic tests using a combined pitot probe system.

3.3.5 Particulate Matter

Particulate matter concentration and emission rate were determined by wet impingement following SCAQMD Method 5.1. A quartz probe with a glass nozzle was used to sample the exhaust stream. Following the probe the sample travels through flexible teflon tubing to the impingers and then to a tared

backup filter. A thermocouple and pitot probe are attached to the probe to facilitate isokinetic sampling at each of the 16 traverse points. Two ports were used to access the traverse points. Duplicate 64-minute samples were collected.

3.3.6 Trace Organic Hydrocarbons

Trace organic species were collected by Carnot in Tedlar bags and analyzed by GC/PID/ELCD by the LACSD lab at JWPCP using the EPA Method T03 and analyzed within 72 hours of sampling. Inlet and exhaust were sampled simultaneously for a twenty-minute period. The flow rate used to calculate the lb/hr emission rates are from the particulate matter test flow rate being performed concurrently with trace organic sampling.

3.3.7 Hydrogen Sulfide and Reduced Sulfur Compounds

Samples for determination of hydrogen sulfide and reduced sulfur compounds were collected in Tedlar bags. The samples were analyzed by GC/FPD/ELCD by the LACSD lab at JWPCP and analyzed within 8 hours of sampling.

SECTION 4.0

RESULTS

The results of the test are presented in the following tables:

Table Number	Table Title
4-1	General Results, PVERG, Unit 1
4-2	Trace Organic Species Destruction Efficiency Results, PVERG, Unit 1, Test No. 1
4-3	Trace Organic Species Destruction Efficiency Results, PVERG, Unit 1, Test No. 2

**TABLE 4-1
GENERAL RESULTS
PVERG, UNIT 1
DECEMBER 30, 1992**

Parameter	INLET			EXHAUST GAS		
	First Run	Second Run	Average	First Run	Second Run	Average
O ₂ , % dry	11.2	11.2	11.2	5.27	5.39	5.33
CO ₂ , % dry	15.8	16.1	16.0	13.01	12.83	12.92
N ₂ , % dry	54.6	54.4	54.5	81.7	81.8	81.8
H ₂ O, %	3.8	4.2	4.0	14.4	14.6	14.5
Flow Rate, cfm	3730	3714	3722	---	---	---
Flow Rate, dscfm	3589	3557	3573	14576	14654	14615
Temperature, °F				336	337	337
NO _x :						
ppm				18.4	18.3	18.4
ppm @ 3% O ₂				21.1	21.1	21.1
lb/hr (as NO _x)				1.95	1.95	1.95
CO:						
ppm				12.0	15.8	13.9
ppm @ 3% O ₂				13.7	18.2	16.0
lb/hr				0.774	1.025	0.900
HC:						
CH ₄ , ppm	192000	199000	195500	ND < 1	ND < 1	ND < 1
TGNMO ppm	7040	7360	7200	15.8	16.3	16.1
TGNMO ppm @ 3% O ₂	--	--	--	18.1	18.8	18.5
TGNMO lb/hr (as CH ₄)	63.92	66.22	65.07	0.583	0.604	0.593
Destruction Eff. %	--	--	--	99.09	99.09	99.09
Particulate:						
Organic Fraction gr/dscf				0.0000	0.0000	0.0000
Inorganic Fraction gr/dscf				0.0034	0.0020	0.0027
Total Particulate gr/dscf				0.0034	0.0020	0.0027
gr/dscf @ 3% O ₂				0.0040	0.0071	0.0056
gr/dscf @ 12% CO ₂				0.0031	0.0019	0.0025
lb/hr				0.424	0.256	0.340
Sulfur Compounds:						
H ₂ S, ppm	21	22	22			
Methyl Mercaptan, ppm	0.5	0.5	0.5			
Ethyl Mercaptan, ppm	ND < 0.5	ND < 0.5	ND < 0.5			
Dimethyl Sulfide, ppm	ND < 0.6	ND < 0.6	ND < 0.6			
Dimethyl Disulfide, ppm	ND < 0.5	ND < 0.5	ND < 0.5			
Carbonyl Sulfide, ppm	ND < 0.4	ND < 0.4	ND < 0.4			
Carbon Disulfide, ppm	ND < 0.6	ND < 0.6	ND < 0.6			
Total Sulfur Compounds, ppm	< 25.2	< 26.2	< 26.2			

TABLE 4-2
TRACE ORGANIC SPECIES DESTRUCTION EFFICIENCY RESULTS
PVERG UNIT 1, TEST NO. 1
DECEMBER 30, 1992

Species	Inlet	Exhaust	Destruction Efficiency, %
Flow rate, Inlet dscfm	3589		
Flow rate, Exhaust dscfm	14576		
Methylene Chloride: lb/hr	200 ppb 9.64×10^{-3}	ND < 0.2 ppb $< 3.91 \times 10^{-5}$	> 99.59
Chloroform: lb/hr	ND < 200 ppb $< 1.35 \times 10^{-2}$	ND < 0.2 ppb $< 5.50 \times 10^{-5}$	NA
1,1,1-trichloroethane: lb/hr	ND < 200 ppb $< 1.51 \times 10^{-2}$	ND < 0.2 ppb $< 6.15 \times 10^{-5}$	NA
Carbon Tetrachloride: lb/hr	ND < 200 ppb $< 1.75 \times 10^{-2}$	ND < 0.2 ppb $< 7.09 \times 10^{-5}$	NA
1,1-dichloroethene: lb/hr	ND < 200 ppb $< 1.10 \times 10^{-2}$	ND < 0.2 ppb $< 4.47 \times 10^{-5}$	NA
Trichloroethylene: lb/hr	500 ppb 3.73×10^{-2}	ND < 0.2 ppb $< 6.06 \times 10^{-5}$	> 99.84
Tetrachloroethylene: lb/hr	400 ppb 3.76×10^{-2}	ND < 0.2 ppb $< 7.64 \times 10^{-5}$	> 99.80
Chlorobenzene: lb/hr	4700 ppb 3.00×10^{-1}	ND < 0.4 ppb $< 1.04 \times 10^{-4}$	> 99.97
Vinyl Chloride: lb/hr	1400 ppb 4.97×10^{-2}	ND < 0.4 ppb $< 5.76 \times 10^{-5}$	> 99.88
m-dichlorobenzene: lb/hr	ND < 200 ppb $< 1.67 \times 10^{-2}$	ND < 0.2 ppb $< 6.78 \times 10^{-5}$	NA
o+p dichlorobenzene: lb/hr	1300 ppb 1.08×10^{-1}	ND < 0.2 ppb $< 6.78 \times 10^{-5}$	> 99.94
1,1-dichloroethane: lb/hr	ND < 200 ppb $< 1.12 \times 10^{-2}$	ND < 0.2 ppb $< 4.56 \times 10^{-5}$	NA
1,2-dichloroethane: lb/hr	ND < 200 ppb $< 1.12 \times 10^{-2}$	ND < 0.2 ppb $< 4.56 \times 10^{-5}$	NA
Benzene: lb/hr	11000 ppb 4.88×10^{-1}	0.2 ppb 3.60×10^{-5}	99.99
Toluene: lb/hr	24000 ppb 1.25×10^0	ND < 1 ppb $< 2.12 \times 10^{-4}$	> 99.98
Acetonitrile: lb/hr	ND < 2000 ppb $< 4.66 \times 10^{-2}$	ND < 2 ppb $< 1.89 \times 10^{-4}$	NA
m-xylenes: lb/hr	25000 ppb 1.51×10^0	ND < 0.2 ppb $< 4.89 \times 10^{-5}$	100.00
o+p xylenes: lb/hr	21000 ppb 1.27×10^0	ND < 0.2 ppb $< 4.89 \times 10^{-5}$	100.0
1,2-dibromoethane: lb/hr	ND < 7000 ppb $< 7.46 \times 10^{-1}$	ND < 7 ppb $< 3.03 \times 10^{-3}$	NA
Benzyl Chloride: lb/hr	ND < 5000 ppb $< 3.59 \times 10^{-1}$	ND < 5 ppb $< 1.46 \times 10^{-3}$	NA

ND < indicates that the species was not detected. Values indicate the detection limit for this species and the concentration is less than the presented value.
 NA - indicates that the destruction efficiency cannot be calculated because the concentration is below detection limits.

TABLE 4-3
TRACE ORGANIC SPECIES DESTRUCTION EFFICIENCY RESULTS
PVERG UNIT 1, TEST NO. 2
DECEMBER 30, 1992

Flow rate, Inlet dscfm	3557			
Flow rate, Exhaust dscfm	14654			
Species	Inlet	Exhaust	Destruction Efficiency, %	
Methylene Chloride: lb/hr	ND < 200 ppb < 9.55×10^{-3}	0.4 ppb 7.87×10^{-5}	< 99.18	
Chloroform: lb/hr	ND < 200 ppb < 1.34×10^{-2}	ND < 0.2 ppb < 5.53×10^{-5}	NA	
1,1,1-trichloroethane: lb/hr	ND < 200 ppb < 1.50×10^{-2}	ND < 0.2 ppb < 6.18×10^{-5}	NA	
Carbon Tetrachloride: lb/hr	ND < 200 ppb < 1.73×10^{-2}	ND < 0.2 ppb < 7.13×10^{-5}	NA	
1,1-dichloroethene: lb/hr	ND < 200 ppb < 1.09×10^{-2}	ND < 0.2 ppb < 4.49×10^{-5}	NA	
Trichloroethylene: lb/hr	400 ppb 2.96×10^{-2}	ND < 0.2 ppb < 6.09×10^{-5}	> 99.79	
Tetrachloroethylene: lb/hr	ND < 200 ppb < 1.87×10^{-2}	ND < 0.2 ppb < 7.68×10^{-5}	NA	
Chlorobenzene: lb/hr	2400 ppb 1.52×10^{-1}	ND < 0.4 ppb < 1.04×10^{-4}	> 99.93	
Vinyl Chloride: lb/hr	1400 ppb 4.92×10^{-2}	ND < 0.4 ppb < 5.79×10^{-5}	> 99.88	
m-dichlorobenzene: lb/hr	ND < 200 ppb < 1.65×10^{-2}	ND < 0.2 ppb < 6.81×10^{-5}	NA	
o+p dichlorobenzene: lb/hr	ND < 200 ppb < 1.65×10^{-2}	ND < 0.2 ppb < 6.81×10^{-5}	NA	
1,1-dichloroethane: lb/hr	ND < 200 ppb < 1.11×10^{-2}	ND < 0.2 ppb < 4.59×10^{-5}	NA	
1,2-dichloroethane: lb/hr	ND < 200 ppb < 1.11×10^{-2}	ND < 0.2 ppb < 4.59×10^{-5}	NA	
Benzene: lb/hr	1100 ppb 4.83×10^{-1}	0.1 ppb 1.81×10^{-5}	100.00	
Toluene: lb/hr	22000 ppb 1.14×10^0	ND < 1 ppb < 2.13×10^{-4}	> 99.98	
Acetonitrile: lb/hr	ND < 2000 ppb < 4.62×10^{-2}	ND < 2 ppb < 1.90×10^{-4}	NA	
m-xylenes: lb/hr	22000 ppb 1.31×10^0	ND < 0.2 ppb < 4.92×10^{-5}	100.0	
o+p xylenes: lb/hr	19000 ppb 1.13×10^0	ND < 0.2 ppb < 4.92×10^{-5}	100.0	
1,2-dibromoethane: lb/hr	ND < 7000 ppb < 7.40×10^{-1}	ND < 7 ppb < 3.05×10^{-3}	NA	
Benzyl Chloride: lb/hr	ND < 5000 ppb < 3.56×10^{-1}	ND < 5 ppb < 1.47×10^{-3}	NA	

ND < indicates that the species was not detected. Values indicate the detection limit for this species and the concentration is less than the presented value.
 NA - indicates that the destruction efficiency cannot be calculated because the concentration is below detection limits.

APPENDIX A
MEASUREMENT PROCEDURES

Continuous Emissions Monitoring System

Oxygen (O₂) by Continuous Analyzer

Carbon Dioxide (CO₂) by Continuous Analyzer

NO/NO_x by Continuous Analyzer

Carbon Monoxide (CO) by NDIR/Gas Filter Correlation

Total Particulate by SCAQMD Wet Impingement Method

Hydrocarbons by SCAQMD Micro TCA Method

Stack Gas Velocity and Volumetric Flow Rate

Determination of Moisture in Stack Gases

Procedure 280A: Gas Chromatographic Analysis of Volatile Sulfur Compounds

Procedure 600B: Gas Chromatographic Analysis of Organic Compounds in Air and Gas Samples
Using a Packed Column

Procedure 600C: Gas Chromatographic Analysis of Acetonitrile in Air and Gas Samples
Using a Packed Column

Procedure 1203A: Gas Chromatographic Analysis of Air, Methane and Carbon Dioxide
Using a Packed Column

Continuous Emissions Monitoring System


O₂, CO, CO₂, NO, NO_x and SO₂ are measured using an extractive continuous emissions monitoring (CEM) package, shown in the following figure. This package is comprised of three basic subsystems. They are: (1) the sample acquisition and conditioning system, (2) the calibration gas system, and (3) the analyzers themselves. This section presents a description of the sampling and calibration systems. Descriptions of the analyzers used in this program and the corresponding reference test methods follow. Information regarding quality assurance information on the system, including calibration routines and system performance data follows.

The sample acquisition and conditioning system contains components to extract a representative sample from the stack or flue, transport the sample to the analyzers, and remove moisture and particulate material from the sample. In addition to performing the tasks above, the system must preserve the measured species and deliver the sample for analysis intact. The sample acquisition system extracts the sample through a stainless steel probe. The probe is insulated or heated as necessary to avoid condensation. If the particulate loading in the stack is high, a sintered stainless steel filter is used on the end of the probe.

Where water soluble NO₂ and/or SO₂ are to be measured, the sample is drawn from the probe through a heated teflon sample line into an on-stack cooled (approximately 35-40°F) water removal trap. The trap consists of stainless steel flasks in a bath of ice and water. This design removes the water vapor by condensation. The contact between the sample and liquid water is minimized and the soluble NO₂ and SO₂ are conserved. This system meets the requirements of EPA Method 20. The sample is then drawn through a teflon transport line, particulate filter, secondary water removal and into the sample pump. The pump is a dual head, diaphragm pump. All sample-wetted components of the pump are stainless steel or teflon. The pressurized sample leaving the pump flows through a third condensate trap in a refrigerated water bath (≈38°F) for final moisture removal. A drain line and valve are provided to constantly expel any condensed moisture from the dryer at this point. After the dryer, the sample is directed into a distribution manifold. Excess sample is vented through a back-pressure regulator, maintaining a constant pressure of 5-6 psig to the analyzer rotameters.

The calibration system is comprised of two parts: the analyzer calibration, and the system bias check (dynamic calibration). The analyzer calibration equipment includes pressurized cylinders of certified span gas. The gases used are, as a minimum, certified to 1% by the manufacturer. Where necessary to comply with reference method requirements EPA Protocol 1 gases are used. The cylinders are equipped with pressure regulators which supply the calibration gas to the analyzers at the same pressure and flow rate as the sample. The selection of zero, span, or sample gas directed to each analyzer is accomplished by operation of the sample/calibration selector fittings.

The system bias check is accomplished by transporting the same gases used to zero and span the analyzers to the sample system as close as practical to the probe inlet. This is done either by attaching the calibration gas supply line to the probe top with flexible tubing or by actuation of a solenoid valve located at the sample conditioner inlet (probe exit). The span gas is exposed to the same elements as the sample and the system response is documented. The analyzer indications for the system calibration check must agree within 5% of the analyzer calibration. Values are adjusted and changes/repairs are made to the system to compensate for any difference in analyzer readings. Specific information on the analytical equipment and test methods used is provided in the following pages.



Schematic of CEM System

Method:	Oxygen (O₂) by Continuous Analyzer
Applicable Reference Methods:	EPA 3A, EPA 20, ARB 100, BA ST-14, SCAQMD 100.1
Principle:	A sample is continuously drawn from the flue gas stream, conditioned, and conveyed to the instrument for direct readout of O ₂ concentration.
Analyzer:	Teledyne Model 326A
Measurement Principle:	Electrochemical cell
Ranges:	0-5, 0-10, 0-25% O ₂
Accuracy:	1% of full scale
Output:	0-100 mV, linear
Interferences:	Halogens and halogenated compounds will cause a positive interference. Acid gases will consume the fuel cell and cause a slow calibration drift.
Response Time:	90% < 7 seconds
Sampling Procedure:	A representative flue gas sample is collected and conditioned using the CEM system described previously. If Method 20 is used, that method's specific procedures for selecting sample points are used. Otherwise, stratification checks are performed at the start of a test program to select single or multiple-point sample locations.
Analytical Procedure:	An electrochemical cell is used to measure O ₂ concentration. Oxygen in the flue gas diffuses through a Teflon membrane and is reduced on the surface of the cathode. A corresponding oxidation occurs at the anode internally, and an electric current is produced that is proportional to the concentration of oxygen. This current is measured and conditioned by the instrument's electronic circuitry to give an output in percent O ₂ by volume.
Special Calibration Procedure:	The measurement cells used with the O ₂ instrument have to be replaced on a regular basis. After extended use, the cell tend to produce a nonlinear response. Therefore, a three-point calibration is performed at the start of each test day to check for linearity. If the response is not linear (\pm 2% of scale), the cell is replaced.

Method: **Carbon Dioxide (CO₂) by Continuous Analyzer**

Applicable Reference Methods: EPA 3A, ARB 100, BA ST-5, SCAQMD 100.1

Principle: A sample is continuously drawn from the flue gas stream, conditioned, and conveyed to the instrument for direct readout of CO₂ concentration.

Analyzer: Horiba PIR 2000

Measurement Principle: Non-dispersive infrared (NDIR)

Accuracy: 1% of full scale

Ranges: 0-5, 0-10, 0-25%

Output: 0-10 mV

Interferences: A possible interference includes water. Since the instrument receives dried sample gas, this interference is not significant.

Response Time: 1.2 seconds

Sampling Procedure: A representative flue gas sample is collected and conditioned using the CEM system described previously.

Analytical Procedure: Carbon dioxide concentrations are measured by short path length non-dispersive infrared analyzers. These instruments measure the differential in infrared energy absorbed from energy beams passed through a reference cell (containing a gas selected to have minimal absorption of infrared energy in the wavelength absorbed by the gas component of interest) and a sample cell through which the sample gas flows continuously. The differential absorption appears as a reading on a scale of 0 to 100%.

Method: **NO/NO_x by Continuous Analyzer**

Applicable Reference Methods: EPA 7E, EPA 20; ARB 100, BA ST-13A, SCAQMD 100.1

Principle: A sample is continuously drawn from the flue gas stream, conditioned, and conveyed to the instrument for direct readout of NO or NO_x.

Analyzer: Teco Model No. 10AR

Measurement Principle: Chemiluminescence

Accuracy: 1% of full scale

Ranges: 0-2.5, 0-10, 0-25, 0-100, 0-250, 0-1000, 0-2500, 0-10,000 ppm

Output: 0-10 mV

Inferences: Compounds containing nitrogen (other than ammonia) may cause interference.

Response Time: 90%, 1.5 seconds (NO mode) and 1.7 seconds (NO_x mode)

Sampling Procedure: A representative flue gas sample is collected and conditioned using the CEM system described previously. If EPA Method 20 is used, that method's specific procedures for selecting sample points are used.

Analytical Procedure: The oxides of nitrogen monitoring instrument is a chemiluminescent nitric oxide analyzer. The operational basis of the instrument is the chemiluminescent reaction of NO and ozone (O₃) to form NO₂ in an excited state. Light emission results when excited NO₂ molecules revert to their ground state. The resulting chemiluminescence is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electronically processed so it is linearly proportional to the NO concentration. The output of the instrument is in ppmV.

When NO₂ is expected to be present in the flue gas, a supercooled water dropout flask will be placed in the sample line to avoid loss of NO₂. Since NO₂ is highly soluble in water, "freezing out" the water will allow the NO₂ to reach the analyzers for analysis. The analyzer measures NO only. In the NO_x mode, the gas is passed through a moly converter which converts NO₂ to NO and a total NO_x measurement is obtained. NO₂ is determined as the difference between NO and NO_x. Use of a moly converter instead of a stainless steel converter eliminates NH₃ interference; NH₃ is converted to NO with a stainless converter, but not with a moly converter.

Method: **Carbon Monoxide (CO) by NDIR/Gas Filter Correlation**

Applicable Reference Methods: EPA 10; ARB 1-100; BA ST-6, SCAQMD 100.1

Principle: A sample is continuously drawn from the flue gas stream, conditioned, and conveyed to the instrument for direct readout of CO concentration.

Analyzer: TECO, Model 48

Measurement Principle: NDIR/Gas Filter Correlation

Precision: 0.1% ppm

Ranges: 0-5, 0-10, 0-25, 0-50, 0-100, 0-250, 0-500, 0-1000, 0-5000 ppm

Output: 0-100 mV

Interferences: Negligible interference from water and CO₂

Rise/Fall Times (0-95%) 1 minute @ 1 lpm flow, 30 second integration time

Sampling Procedure: A representative flue gas sample is collected and conditioned using the CEM system described previously. Sample point selection has been described previously.


Analytical Procedure: Radiation from an infrared source is chopped and then passed through a gas filter which alternates between CO and N₂ due to rotation of a filter wheel. The radiation then passes through a narrow band-pass filter and a multiple optical pass sample cell where absorption by the sample gas occurs. The IR radiation exits the sample cell and falls on a solid state IR detector.

Method: **Total Particulate by SCAQMD Wet Impingement Method**

Reference: SCAQMD Method 5.1

Principle: A metered flue gas sample is withdrawn isokinetically and particulates are collected in a series of water impingers followed by a backup filter.

Sampling Procedure: The sample train used in the tests is shown in the following figure. The sample is drawn isokinetically through a stainless steel or glass nozzle and probe. The probe is connected to an impinger train by Teflon tubing. The train consists of two Greenburg-Smith impingers which contain 100 ml of distilled water, an empty impinger as a knockout, a 100 mm glass fiber filter, and an impinger containing silica-gel to protect the leak-tight vacuum pump and calibrated dry gas meter from moisture. The impingers are designed to collect particles larger than one micron and will also collect some smaller water-soluble particles. The backup filter collects any particulate that is not collected in the impingers. This material collected in the backup filter consists of submicron-sized droplets of condensed organic vapors.



Sample Train for Determination of Total Particulate by SCAQMD Wet Impingement Method

SCAQMD Method 4.1 (moisture) and Methods 1.1 and 2.1 (velocity) are performed in conjunction with the test. Stack velocity is measured during the test to maintain isokinetic sampling and to measure stack flow rate. Moisture concentration is determined by weighing the impingers before and after sampling to determine the amount of moisture collected.

Sample Recovery and
Analysis:

Following testing, the probe, sample line, and impingers are washed with water. The entire rinse and sample is then filtered through a tared filter to recover large particles. An aliquot of the rinse is evaporated in a tared glass beaker at 105°C, and the residue is determined gravimetrically. If the sample contains semi-volatile hydrocarbon species that might be lost by evaporating at 105°C, the sample is extracted with methylene chloride to remove those species, and the methylene chloride fraction is evaporated at ambient temperature and weighed.

The backup filter is baked at 105°C, desiccated, cooled, and weighed, and the weight combined with the boildown and other filter to determine the total particulate catch.


If there are high concentrations of SO₂ in the gas stream, a positive interference can be caused by oxidation of SO₂ to sulfuric acid in the impingers. In this case, the boildown residue is analyzed for sulfuric acid (either by titration or precipitation) and the acid is subtracted from the particulate catch as H₂SO₄·2H₂O. Since this fraction includes both SO₂-related "pseudo-particulate" as well as the true particulate species SO₃, a separate test for SO₃ is run and the measured SO₃ is added back to the particulate catch as H₂SO₄·2H₂O.

Method: **Hydrocarbons by SCAQMD Micro Total Carbon Analysis (TCA/FID) Method**

Reference: SCAQMD Method 25.2 (draft Source Test Manual)

Principle: A Tedlar bag is filled with flue gas at a constant rate. The bag contents are analyzed by gas chromatography/flame ionization detection for CO, CO₂, methane and total gaseous non-methane hydrocarbons.

Sampling Procedure: Samples are collected using a lung-type sampling system shown in the attached figure. In this system, a bag is placed in a sealed container and the container is evacuated. Flue gas enters the bag as it expands to fill the container. Sampling rate is monitored by a rotameter on the container exhaust. This system allows sample collection without exposing the sample to pumps, flowmeters, oils, etc.



**Sample Train for Determination of Hydrocarbons
by SCAQMD Micro TCA Method**

Analytical Procedure:

In the analytical phase, gaseous carbon compounds from the tank are fractionated on a chromatographic column, eluting in the order: carbon monoxide, methane, carbon dioxide, non-methane hydrocarbons. All resulting vapors are passed through oxidizers where they are combusted to carbon dioxide, then converted to methane and measured by flame ionization detection. CO₂, CO, CH₄, and total gaseous non-methane hydrocarbons are reported separately.

This method provides lower detection limits than the standard TCA procedure, which uses GC/NDIR. Since this method does not use a steel condensate trap, it eliminates positive interferences that have been associated with carbon dioxide dissolved in water frozen in the condensate trap and released during analysis.

The TCA/FID method is appropriate for combustion sources and ambient sampling where low concentrations of hydrocarbons are expected.

Carnot subcontracts TCA analysis to qualified local laboratories experienced in the analytical procedures.

Method: **Stack Gas Velocity and Volumetric Flow Rate**

Reference: EPA Method 2, SCAQMD Method 2.1, ARB Method 2

Principle: The average gas velocity in a stack is determined from the measurement of the gas density and from the measurement of the average velocity head using a Type-S (Stausscheibe) Pitot tube.

Sampling Procedure: The velocity head and temperature are measured at traverse points specified by EPA Method 1 or SCAQMD Method 1.1. The velocity is measured using a Type-S Pitot tube and an inclined water manometer. The flow coefficient of the pitot tube is known. Temperature of the gas is measured using a thermocouple. The stack gas molecular weight is determined from independent measurements of O₂, CO₂, and H₂O concentrations.

Sample Analysis and Recovery: The stack gas velocity is determined from the measured average velocity head, the measured average temperature, the measured average duct static pressure, the measured dry concentrations of O₂ and CO₂, and the measured concentration of H₂O. The velocity is determined from the following set of equations:

$$V_s = 2.90 C_p \sqrt{\Delta p T_s \left[\frac{29.92}{P_s} \right] \left[\frac{28.95}{MW_{wet}} \right]} \quad [\text{ft/s}]$$

$$\Delta p = \text{Velocity/Head, inches } H_2O \quad [\text{in. } H_2O]$$

$$T_s = \text{Gas Temperature, degrees R} \quad [\text{R}]$$

$$P_s = \text{Absolute Static Pressure} \quad [\text{in Hg}]$$

$$C_p = \text{Pitot Flow Coefficient} \quad [\text{unitless}]$$

$$MW_{wet} = [(0.44)(\%CO_2) + (0.32)(\%O_2) + (0.28)(\%N_2)] \left(1 - \frac{\%H_2O}{100} \right) + (18) \left(\frac{\%H_2O}{100} \right)$$

The stack gas volumetric flow rate is determined from the measured stack gas velocity, the area of the stack at the measurement plane, and the measured gas temperature and pressure. The volumetric flow rate is determined from the following set of equations:

$$Q = (V_s)(AREA)(60) \quad [\text{wacfm}]$$

$$Q_{ws} = Q \left[\frac{T_{ref}}{T_s} \right] \left[\frac{P_s}{29.92} \right] \quad [\text{wscfm}]$$

$$Q_{sd} = Q_{ws} \left[1 - \frac{\%H_2O}{100} \right] \quad [\text{dscfm}]$$

Method:

Determination of Moisture in Stack Gases

Applicable Ref. Methods:


EPA 4, ARB 1-4, SCAQMD 4.1

Principle:

A gas sample is extracted at a constant rate from the source; moisture is removed from the sample stream and determined volumetrically or gravimetrically.

Sampling Procedure:

The sample train used in the tests is shown in the following figure. The sample is drawn at a constant rate through a stainless steel probe. The probe is connected to an impinger train by Teflon tubing. The train consists of two Greenburg-Smith impingers which contain 100 ml water, an empty impinger as a knockout, and an impinger containing silica gel to protect the pump from moisture.



Sample Train for Determination of Moisture by EPA Method 4

Sample Recovery and Analysis:

Following testing, moisture content is determined gravimetrically from initial and final impinger weights.

280A. GAS CHROMATOGRAPHIC ANALYSIS OF VOLATILE SULFUR COMPOUNDS

INTRODUCTION

Gaseous sulfur compounds are cryogenically collected on a solid phase sorbent trap and analyzed by packed column gas chromatography. The gas chromatograph is equipped with two different columns and two different detectors operating in parallel in order to effect positive identification and quantitation of the target analytes.

Procedure 280A: Gas Chromatographic Analysis of Volatile Sulfur Compounds (12-31-91)

1. Scope and Application

- 1.1 This procedure is used for the identification and quantitation of the sulfur gases listed in Table I. Quantitation may be carried out at the ppb level under proper conditions.
- 1.2 Samples encountered include landfill, digester and sewer gas and ambient air. This analytical method requires a 30 minute chromatographic run and is used for ppb-level analysis of a wide range of sulfur compounds. An abbreviated method which measures the target compound list of SCAQMD Rule 1150.1 at the part-per-million level requires a 15 minute run.

2. Summary of Procedure

- 2.1 Known amounts of sample gas are passed through liquid nitrogen-cooled traps. The traps are swept with helium to remove oxygen and nitrogen and are then heated to desorb the trapped sulfur gases onto cooled (-60°C) gas chromatographic columns. The columns are then heated at a rate which allows these compounds to be separated on the respective columns. The two detectors, a flame photometric detector (FPD) and a Hall electrolytic conductivity detector (HECD), may be interchanged if further confirmatory analyses are necessary.

3. Sample Handling and Preservation

- 3.1 Samples are normally collected in Tedlar (polyvinyl fluoride) bags and brought to the laboratory for analysis. Samples should be stored in the dark and away from heat. Samples should be analyzed within twenty-four hours. This is especially important when high concentrations (1 ppm) of methyl mercaptan are suspected.

4. Interferences

- 4.1 The flame photometric detector operates by detecting the fluorescence of S_2 molecules formed from sulfur atoms which are liberated when sulfur-containing species are burned in a hydrogen/air flame. Since the fluorescence intensity is directly proportional to the concentration of S_2 molecules and to the square of the sulfur-atom concentration in the flame, the detector signal will be roughly

proportional to the square of the amount of sulfur-containing species reaching the detector. For this to be true, equilibrium must be rapidly established in the flame between sulfur atoms and S_2 molecules. This will occur if the concentration of sulfur-containing species is small compared to the concentration of hydrogen and oxygen. If the detector is overloaded, a chemical saturation of the detector will occur and there will be a loss of sensitivity. This chemical saturation will normally occur at a higher concentration than is required to saturate the photomultiplier and so should not be a problem.

- 4.2 The FPD has a very large rejection ratio for non-sulfur-containing species, so there are very few interferences.
- 4.3 Ethyl mercaptan and dimethyl sulfide co-elute on the Chromosil 330 column under the conditions used in this procedure, but they are completely separated on the XE-60 column.
- 4.4 The sulfur mode HECD reacts sulfur-containing compounds with an excess of oxygen to form SO_2/SO_3 , which is then detected in the conductivity cell. The response due to any CO_2 formed under these oxidizing conditions is levelled by employing a non-aqueous solvent (reagent methanol) to prevent the formation of carbonic acid. Since hydrogen halides will also be formed even under oxidizing conditions, the sulfur mode HECD employs a copper tube containing a silver wire between the reactor and detector as a hydrogen halide scrubber to eliminate interferences from halogenated compounds. This scrubber is not so effective in removing hydrogen fluoride as it is in removing the other halides, and so interferences due to fluorinated hydrocarbons can occur, especially in the region where the C3 and C4-mercaptans elute. Sewer gas typically shows several large peaks in this region which are not due to sulfur containing compounds.
- 4.5 Since the Chromosil 330 column has an upper temperature limit of $100^\circ C$, the XE-60 column cannot be used close to its upper temperature limit. Under these conditions, the retention times of the C3- and higher mercaptans on the XE-60 may be unacceptably long, and the usefulness of the XE-60 is limited to the quantitation of the C1- and C2-mercaptans and sulfides of the SCAQMD rule 1150.1 list.

5. Apparatus

- 5.1 Varian Model 3400 gas chromatograph equipped with a Varian dual-flame FPD and a Tracor Model 700A HECD operated in the sulfur mode.
 - 5.1.1 Columns:
 - a. 8'(6' packed) x 1/8"OD Teflon (FEP) packed with Chromosil 330.
 - b. 6' x 1/4"OD (2 mm id) glass packed with 1.5% XE-60/1.0% H_3PO_4 on Carbopak B. (The glass column must be deactivated before packing by baking at $130^\circ C$ for six hours with a carrier gas flow rate of 100-200 mL/min.)

5.1.2 Operating Parameters

Carrier gas: Helium (carrier grade), 25 mL/min (column a), 30 mL/min (column b).

Detector temperature: 180°C

FPD Range: 10×10^{-8}

PMT voltage: 480-520 volts

HECD Range: 100

HECD Solvent: 100% methanol (reagent grade)

Temperature program: 3.5 minute hold at initial temperature of -60°C. Program to 40°C at 40° C/minute and hold for 4.5 minutes. Program to 65° C at 40°C per minute and hold for 3.0 minutes. Program to 95°C at 40°C/min and hold for 15 minutes. Vent closes at 4.5 min. and opens at 14.75 min.

5.2 Data System: Nelson Analytical (3000 Series) chromatography software installed on an IBM PS/2 with a two-channel buffer box (760 Series) to collect raw data from both detectors simultaneously.

5.3 Cryo-adsorbent Trapping System

5.3.1 Plumbing: 1/8" or 1/16" Teflon tubing used throughout.

5.3.2 Trap design: Two similar U-shaped traps are constructed of 1/8" OD Teflon-lined stainless steel (1.8 mm ID) and packed with 0.1 gram 50/80 mesh Porapak QS. The jacket consists of 1/4" OD copper tubing with Swagelock TEE fittings at the ends. The jacket is wrapped with 12' x 3/16" nichrome heating cord controlled by a Cole-Parmer model 2149-00 temperature control relay. The trap is heated to approximately 210°C during the desorption cycle. The trap is cooled by flowing liquid nitrogen through its outer jacket.

5.4 Sample Loading:

a) small volume samples (1-50 mL) are injected by gas-tight syringe directly into the sweep helium and are deposited on the trap.

b) large volume samples are withdrawn from sample bags with 100 mL glass syringe with Teflon plunger and forced through trap.

5.5 Sample Injection: Samples are transferred from the two traps and injected onto the two columns using twin Valco E60 two-position electric actuator/valve assemblies. Valves are actuated by relays controlled by the GC method which injects at 0.5 min and returns to load position at 10 min. Valves are fabricated from Hastelloy C in order to minimize sulfur gas adsorption and reactivity.

5.5.1 Pneumatics and valves. Samples are introduced through the syringe or through a septum port. The sample gas flows through the trap switch valve into the proper injection valve and liquid nitrogen-cooled trap. With the trap

switch valve in the LOAD A position, the sample is directed to injection valve A and the sweep helium flows through injection valve B. When this valve is in the LOAD B position, sample gas flows to valve B and sweep helium to valve A. Sweep helium can be directed to both injection valves and traps simultaneously by removing the sample needle from the sample bag and inserting it in the sweep helium septum port.

When the sample gas reaches the proper injection valve, it is directed to the liquid nitrogen-cooled trap where the less-volatile species are adsorbed on the Porapak QS. The sweep helium then removes most oxygen, nitrogen and other permanent gases and sends them out the vent. When the valve is turned to the INJECT position, the carrier helium flow is directed to backflush the trap onto the gas chromatographic column and the sweep helium is vented. The two injection valves are configured identically.

After the separated analytes exit the chromatographic columns, they are directed to the detector switch valve. This is a four-port Hastelloy C valve which allows column effluent to be directed to either of the two detectors. Normal configuration has Chromosil 330 effluent going to the flame photometric detector and XE-60 effluent to the HECD.

5.6 Electrical connections. Electrical connections are made between the Varian 3400, the Nelson 760 buffer box, the IBM PS/2, and the cryo-adsorbent system.

The 760 interface is the heart of the system, as it provides the interface between the other members. The connections to the 760 are shown. J101 is the connection to the IBM PC/AT. J1 is the 14-pin connector carrying the relays which control the injection valves, the liquid nitrogen flow to the traps and the trap heaters. Only relays 2 through 7 are used; relays 1 and 8 are reserved by Nelson for use with autosamplers. DET A +/- and DET B +/- are the inputs from the FPD and HECD, respectively. The FPD signal is taken from the side port on the Varian 3400 and HECD signal is taken from the 10VFS and COMP COM terminals on the back of the Tracor 700A module. The START NO IN/COM terminals are connected to J23 in the Varian 3400. This terminal provides a one-second contact closure when the START button in the GC is pushed and is used to provide a "start" signal to the Nelson system. Unused terminals on the Nelson 760 box include the J2 BCD connector and the READY IN/COM connections.

Relay 2 activates the liquid nitrogen flow to the traps and can be manually deactivated at the toggle switch mounted below the traps. Relay 3 controls the trap heaters and can also be manually deactivated by a Cole-Parmer 2149 temperature control relay. Relays 4 and 5 control the position of injection valve A. Relay 4 is activated to move the valve to the INJECT position and relay 5 is activated to return the valve to the LOAD position. Relays 6 and 7 serve the analogous functions on injection valve B. The relays are normally controlled by the Nelson software.

Output from the detectors is also directed to a strip chart recorder in order to provide real-time monitoring of the chromatographic runs.

5.7 Safety Precautions

Liquid nitrogen is used as a coolant for both the adsorbent traps and the column oven. Caution must be exercised due to the possibility of freeze burns caused by exposure of skin to extremely low temperatures. Avoid contact with any metal tubing through which liquid nitrogen is flowing and make sure that the exhaust vents on the adsorbent traps are pointing away from anyone in the laboratory.

Gaseous nitrogen is also an asphyxiant. Since the adsorbent traps use large quantities of liquid nitrogen during each run, the possibility exists that room oxygen could be depleted if the ventilation system fails. Do not operate the system if the laboratory ventilation system is not functioning properly.

When liquified gases vaporize, they expand to approximately 700 times their liquid volumes. Thus, the very real possibility exists for a violent rupture of the liquid nitrogen line if liquid nitrogen is confined between two valves with no access to a relief valve. Make absolutely sure that there is a relief valve between any two closures or valves. **ANY SECTION OF ISOLATED LINE MUST BE VENTED OR HAVE A PRESSURE RELIEF VALVE!**

6. Standards

- 6.1 Hydrogen sulfide, carbonyl sulfide, methyl mercaptan, ethyl mercaptan, carbon disulfide and dimethyl sulfide are obtained from Scott Specialty Gas cylinders in 100 ppm v/v (nominal) concentrations. A second cylinder containing approximately 5 ppm v/v H₂S and 2 ppm v/v each of the other above-named species except ethyl mercaptan is used for daily calibrations. Sulfur dioxide, ethyl mercaptan, and DMDS are taken from separate cylinders containing 5 ppm v/v SO₂, 2 ppm ethyl mercaptan, and 4.6 ppm v/v DMDS in nitrogen.

A stock solution of 0.5% v/v (in methanol) of the heavier components (C3 and higher mercaptans and sulfides) is used to make up the rest of the standards, if needed. The stock solution is prepared by using a 0.05 mL syringe to place 0.05 mL of each of the neat compounds in a 10 mL volumetric flask partially filled with methanol and then diluting to 10 mL. These compounds are obtained from Aldrich (typically 98-99% purity) and are used with no further purification.

- 6.2 Preparation of standards. The Scott 5 ppm/2 ppm v/v standards are taken directly from the cylinder with no further preparation. Two ppm bag standards are made from 100 ppm v/v Scott cylinders by injecting 40 mL of the gas into a Tedlar bag filled with 1960 mL of nitrogen. Bag standards of the heavier components are made by injecting 1 microliter of the 0.5% liquid standard into a Tedlar bag filled with 1 liter of nitrogen. This will give the following concentrations: C3 mercaptans, 1.3 ppm; C4 mercaptans, 1.1 ppm; dimethyl disulfide, 1.4 ppm; and thiophene (if used), 2.2 ppm.

7. Procedure

- 7.1 Load method CROM33AD into the Nelson box.

- 7.9 Calibration curves. Six-point calibration data are generated by injecting 0.5, 1, 2, 5, 10 or 15 mL of the Scott 5 ppm/2 ppm v/v standard on the FPD side, and 0.2, 1, 2, 5, 10, or 15 ml on the HECD side. The C3-C4 mercaptan standard is injected on the Chromosil 330 side only, due to the unacceptably long retention times of these compounds on the XE-60 column. Since dimethyl sulfide and ethyl mercaptan co-elute on the Chromosil 330 column, ethyl mercaptan is injected only on the XE-60 column. The areas and retention times are stored under MMDDFVV for FPD (VV=volume injected) and MMDDHVV for HECD.

Calibration curves are generated by entering peak area counts versus component concentration in the appropriate recalculation method. Calibration data are generated by using a varying injection volume of constant concentration instead of a constant volume of varying concentration, so a universal unit of ppm-mL is used in the recalculation methods instead of a concentration. Differences in injection volume are accounted for in the recalculation methods. The 15 mL injection of hydrogen sulfide is not used on the FPD calibration due to poor fit on the curve.

- 7.10 Updating calibration curves. Calibration curves are updated by averaging the current peak areas with the existing peak areas. This procedure weights the updated calibration point heavily in favor of the current day's response, but does not ignore the past history of the instrument response. The new calibration points should agree to within +/- 20% of the current point since these should not be taken until after "normal" response has been verified. If there is a gross discrepancy, a faulty injection should be suspected.
- 7.11 Recalculation methods. The Nelson method CROM33AD is used to control the relays and data acquisition. Data are stored in files named CHROMxxx (FPD data) and DHROMxxx (HECD data). Recalculation methods for these data are MMFPDDD and MMSHALDD, respectively, which contain the calibration curves for each column/detector combination.
- 7.12 The calibration curves cover a concentration range of greater than one order of magnitude. There are many samples which will generate analyte peaks which are either above or below the range of peak areas covered by the curves. In these cases it may be necessary to perform a second analysis using a larger or smaller injection volume, as appropriate.
- 7.13 Each chromatographic run must be reviewed soon after its completion in order to detect any range, interference or instrumental problems which would require a repeat analysis. Next-day review is not acceptable due to the possibility of sample degradation. Although the stability limits of the target compounds have been established when they are at low concentration in pure air or nitrogen, the complex nature of the samples usually received precludes the assumption of their long-term stability.

8. Calculations

- 8.1 The flame photometric detector response is inherently nonlinear, and calibration data

are fitted to a quadratic function of the form

$$A = a(X)^2 + b(X) + c$$

where A is the peak area, (X) is the analyte concentration in ppm-mL units, and a, b, and c are coefficients generated by the Nelson curve-fitting algorithm. To find the actual sample concentration, Nelson solves for X and divides this value by the injection volume:

$$\text{Concentration} = (X)/V$$

where V is the injection volume.

The Nelson software should find the analyte concentration by solving the quadratic equation when given the analyte peak area and injection volume. However, several versions of the Nelson software are known to contain bugs which prevent the quadratic from being solved correctly. It is then necessary to manually perform the required calculations.

- 8.2 The Hall detector calibration data is plotted on a point to point curve. At higher concentrations, the hydrogen sulfide data is better fitted on a cubic fit equation.
- 8.3 Detection limits have been established for the Chromosil 330/FPD and XE-60/HECD combinations according to the EPA approved method of Glaser, et.al. Most of the target analytes containing one sulfur atom have a detection limit of about 2.5 ppb (assuming a 500 mL injection) on the Chromosil 330/FPD. Compounds such as carbon disulfide and dimethyl disulfide which have two sulfur atoms have somewhat lower limits. Detection limits on the XE-60/HECD have not been completely determined; however it is known that they are well below one part-per-billion (again assuming a 500 mL injection) for all target analytes. If a 5 mL injection volume is used the reporting limits will be <0.1 ppm for the aforementioned areas. Any components that are below the lowest calibration point are reported as less than whatever that calibration point is. Because the XE-60/HECD combination is more sensitive and it separates all 1150.1 compounds, the reported detection limits for the 1150.1 compounds are usually based on this detector/column combination.
- 8.4 Equations used in standards preparation. Concentrations of Tedlar bag standards prepared from Scott cylinders may be calculated from the equation

$$C(\text{bag}) = C(\text{Scott}) \times V(s)/V(\text{tot})$$

Where C(bag) is the concentration of the standard, C(Scott) is the concentration of the relevant component in the Scott cylinder, V(s) is the volume of the standard injected into the bag, and V(tot) is the total volume of standard and diluent in the bag.

Concentrations of standards prepared from the 0.5% liquid standards may be calculated from the equation

$$C(\text{bag}) = \frac{V \times d \times 10 \times 22.4 \text{ L/mole} \times C(\text{comp})}{M_w \times V(\text{bag})}$$

where V is the volume of liquid standard in microliters injected into the Tedlar bag, d is the density of the relevant component in g/mL, C(comp) is the concentration of a given component of the liquid standard in volume percent, 22.4 L/mole is the molar volume of a gas at room temperature, M_w is the molecular weight of the component in amu, and V(bag) is the volume of diluent gas in liters. If C(comp) is assumed to be 0.5%, this equation reduces to

$$C(\text{bag}) = \frac{V \times d \times 112}{M_w \times V(\text{bag})}$$

Note that if a new liquid standard is prepared with a component concentration other than 0.5%, the constant term in this equation will have to be multiplied by a factor of

$$x/0.5 \quad (=2x)$$

where x is the per cent concentration of a given component in the new standard.

9. Quality Assurance Guidelines

- 9.1 A duplicate and spike will be run every tenth sample.
- 9.2 AUTOSQC. The area tables should be entered under FPDSGC and HECDSGC. The retention times are entered under FPD-RT and HECDS-RT. For each detector, a "warning situation" exists whenever any component is out by two standard deviations. A "out of control" situation exists whenever any component is out by 3 standard deviations. Since two detectors are used, you may still be able to do the analysis using the other detector, if it is in control.

If a "warning" or "out of control" situation occurs, then re-inject the standard. If a "out of control" situation repeats, then investigate for possible causes of the problem. If the problem cannot be resolved, then the detector response must be re-calibrated. If a "warning" situation repeats, then investigate for possible causes of the problem. If the problem cannot be resolved, then the samples may still be run. If a "warning" situation persists for three days, then it is considered "out of control".

After the problem has been resolved, then delete all "out of control" points from the control charts. Do not remove the "warning" points.

- 9.3 Spikes for qualitative analysis.
- 9.3.1 Spikes are employed to confirm the identity of the higher molecular weight analytes eluting on the Chromosil 330/FPD combination. Many interfering compounds have retention times very close to those of the target analytes, and positive identification cannot be made on the basis of retention time

alone. The low molecular weight analytes are usually free of interferences with the exception of carbonyl sulfide. Spikes are sometimes necessary to confirm its identity on the Chromosil 330/FPD side. In some samples, a large concentration of a condensable gas will give rise to pressure surges which will shift the retention times of the early eluting peaks. A spike will be necessary to confirm their identities in this case.

Since the target compounds can exist in any combination, concentration range and/or ratio, it is impractical to make up a spiking mixture corresponding to each sample. A standard spiking mixture can be made up which contains all of the standard compounds. "Overspiking" (where a standard is added to a sample that has previously been loaded on a cryogenic trap) with compounds not in the sample is desirable as they can be used to determine whether the spike recovery is quantitative.

Use a quantity of spike approximately equal to the amount of the target compound measured. If there is a wide range of apparent analyte concentrations, use a quantity approximately midway between the highest and lowest concentrations of the target analytes.

- 9.4 Analyte identification. The primary identifying characteristic of the target analytes is their retention time on each of the columns. These are listed in Table I. Note that the retention times will vary somewhat. While the C1-C2 mercaptans elute with very few interfering compounds, the heavier compounds normally elute in a region with many interfering peaks. "Overspiking" is usually necessary for positive identification of these. In some cases, it may be advantageous to switch detectors so that Chromosil 330 effluent goes to the HECD and XE-60 effluent goes to the FPD. Since the two detectors operate on different physical principles and have different interferences, a positive response on both detectors can be taken as a positive confirmation. Detector switching may be useful when a dirty sample is being run which gives many peaks in the region of the target analyte and an overspike does not unambiguously determine peak identity.

The lighter compounds are normally confirmed by the observation of a positive response on both column/ detector combinations. The calculated concentrations of the analyte on both detectors should agree reasonably well. However, because dimethyl sulfide co-elutes with ethyl mercaptan on the Chromosil 330 column, the calculated concentrations of dimethyl sulfide may not agree on the two detectors. If the calculated concentrations do not agree for any of the other components, an interference should be suspected and an overspike may be necessary.

In some instances, a positive response will be seen on one column and not the other. Since at very low concentration levels, the HECD is more sensitive than the FPD, it is possible that a compound could be observed in the XE-60/HECD combination but not on the Chromosil 330/FPD. This will usually occur at the sub-ppb V/V level. Since the heavier compounds are normally measured on the Chromosil 330/FPD only, an overspike is used to provide confirmation rather than relying on dual-column observation.

10. Precision and Accuracy

- 10.1 No precision or accuracy data are currently available in the literature.
- 10.2 Repetitive injections of a known standard on the Chromosil 330/FPD side over a two month period indicate that long-term reproducibility on the FPD is approximately 20%. Back-to-back injections of standards are normally reproducible to 1-2%.
- 10.3 This analytical method was developed at the JWPCP Water Quality Laboratory.

11. References

1. Glaser, J.A., Foerst, D.L., McKee, G.D., Quave, S.A., and Budde, W.L., *Environ. Sci. Technol.*, 1981, 15, 1426.
2. ACS Subcommittee on Environmental Analytical Chemistry, *Anal. Chem.*, 1980, 52, 2242.

TABLE I

Volatile Sulfur Compounds Analyzed by Gas Chromatography

Chromosil 330/FPD	RT(min)	XE-60/HECD	RT(min)
carbonyl sulfide	6.04	hydrogen sulfide	5.31
hydrogen sulfide	6.29	carbonyl sulfide	6.27
methyl mercaptan	7.74	sulfur dioxide	7.06
carbon disulfide	8.29	methyl mercaptan	7.22
sulfur dioxide	10.0	ethyl mercaptan	11.02
dimethyl sulfide	9.6	dimethyl sulfide	11.34
ethyl mercaptan	9.6	carbon disulfide	11.87
iso-propyl mercaptan	10.85		
n-propyl mercaptan	12.85		
t-butyl mercaptan	13.26		
thiophene	14.6		
sec-butyl mercaptan	15.0		
iso-butyl mercaptan	15.4		
n-butyl mercaptan	16.0		
dimethyl disulfide	16.5		

1203A: GAS CHROMATOGRAPHIC ANALYSIS OF AIR, METHANE AND CARBON
 DIOXIDE USING A PACKED COLUMN

INTRODUCTION

Permanent gases are relatively stable gases (oxygen, argon, nitrogen, carbon dioxide, methane and carbon monoxide) present in air. The relative amounts of these light gases provide us with important information regarding aerobic and anaerobic operations at District facilities and nearby areas. Additionally, permanent gas analyses can be used as screens to give laboratory personnel an approximate volume to inject on other gas chromatographs when trace levels of volatile organics are to be analyzed. Samples are analyzed for permanent gases by packed column chromatography utilizing thermal conductivity detection (TCD). The detector signals are collected by a data system for identification and quantitation of the sample components.

Procedure 1203A: Gas Chromatographic Analysis of Air, Methane, and Carbon Dioxide Using
 a Packed Column (12-26-91)

1. Scope and Application

This instrumental procedure is to be used for the identification and quantitation of air, methane, and carbon dioxide. Digester gas, engine exhaust, flare exhaust, and landfill gas have been analyzed using this procedure.

2. Summary of Procedure

- 2.1 The sample is injected directly onto a packed column with a 1.0 mL gas-tight syringe.
- 2.2 The gas chromatograph is operated isothermally to separate the components and TCD is used for component detection.

3. Sample Handling

- 3.1 Samples are collected in the field in either glass bulbs or aluminized, Mylar-coated bags equipped with screw valves or Tedlar (polyvinylfluoride) bags equipped with screw valves.
- 3.2 Glass bulbs are flushed with laboratory air prior to sampling and are then evacuated. One end of the bulb is connected to the sampling port and both stopcocks are opened. The sample is allowed to purge the bulb for 30 to 60 sec. The stopcocks are then closed and the sample is transported to the laboratory for analysis.
- 3.3 Aluminized or Tedlar bags are evacuated in the laboratory prior to use. The screw valve is opened and connected to the sampling port. The bag is then allowed to fill under pressure. The valve is closed and disconnected from the sampling port.
- 3.4 Samples are to be analyzed within 48 h.

4. Apparatus

4.1 Gas Chromatograph

4.1.1 Varian 3400 Series gas chromatograph equipped with thermal conductivity detectors.

4.1.2 Packed column: 2m x 2mm Pyrex packed with 80/100 mesh Porapak QS (Alltech Associates, Inc.).

4.2 Data System

4.1.1 Nelson Analytical (3000 series) chromatography software system based on an IBM AT or PS/2 personal computer with a two channel capacity buffer box (760 series) to collect the raw data from the TCD. The data system draws baselines and identifies components according to the software and preselected parameters. The data may be recalculated by the operator by changing the drawn baselines.

4.1.2 Permgas.Bas is a basic program written using the BASIC interpreter. This program is designed to calculate the oxygen and argon concentrations using the O₂-Ar concentration obtained from the Nelson data system. See 7.3 and 7.4. It also averages the methane results from the two permanent gas methods and sums the components for a total permanent gas concentration.

4.3 Syringes

One mL glass Pressure-Lok Series A-2 syringes equipped with a Teflon plunger and push-button valve (Precision Scientific) are used.

5. Procedure

5.1 Set the gas chromatograph to the following parameters:

5.1.1 Carrier grade Helium (99.9995% purity) at 80 mL/min.

5.1.2 Injector temperature: 110 C.

5.1.3 Column temperature: 60 C isothermal.

5.1.4 Detector temperature: 110 C.

5.2 Set the thermal conductivity detector as follows:

5.2.1 Carrier grade Helium (99.9995% purity) reference gas at 80 mL/min.

5.2.2 Filament temperature: 130 C.

- 5.2.3 Filament current: 220 mA.
- 5.2.4 Detector polarity: Positive.
- 5.2.5 Detector initial range: 0.5
- 5.5.6 Detector initial attenuation: 8
- 5.5.7 TCD autozero on: YES
- 5.3 Ready the data system for Nelson data collection prior to the start of the run.
- 5.4 Inject 1.0 mL of standard or sample onto the column of the gas chromatograph.
- 5.5 After a 2.0 min analysis, analytes are quantitated by the data system.

6. Standards

- 6.1 A standard calibration is performed each day a sample is run. Linearity studies indicate that the TCD response to these compounds is quite linear over the concentration ranges shown in Table 1. All samples currently analyzed using this method fall within these concentration ranges.
- 6.2 An air blank is run each day before the standard. This blank is used to determine the amount of contamination of air due to the syringe technique and to insure that the standard line is not contaminated with air. The standard line should be purged for 30 seconds. The pressure in the line should then be adjusted to between 1 and 5 PSI. One mL of the CH₄/CO₂ standard is then drawn and injected. The area count for Air must be below 1,000 before proceeding. This blank value is entered into the calibration curve.
- 6.3 The syringe is filled with 0.50 mL of a 63.8% CH₄ and 36.3% CO₂ mix (Scott Specialty Gases), and it is then diluted with 0.50 mL of desiccated laboratory air. The air is desiccated by drawing laboratory air through a glass tube filled with silica gel desiccant. This mix is then injected as a standard. If the results obtained are 97% to 103% of the theoretical values (as compared to the previous calibration standard), the standard run is accepted for the calibration standard. If the results are not within this range, two or three standard injections are performed. If the results still fall out of range, trouble-shooting procedures are initiated.

7. Calculations

- 7.1 The data system identifies and quantifies sample components
- 7.2 Two-point calibration curves (point-to-point) are established for each component using the Nelson data system.

7.3 Air, methane, and carbon dioxide are calculated by the data system as follows:

$$\text{amount of component (\%)} = \frac{(\text{std amt (\%)} - \text{blank amt})}{(\text{std area} - \text{blank area})} \times \text{sample area in sample}$$

The blank value is determined for air only. Blank areas and amounts are assumed to be zero for methane and carbon dioxide.

7.4 Values below the method detection limit (MDL) are reported as less than the MDL value. See Table 1. Three significant figures are used for reporting data. Figures past the decimal place of the MDL are not used.

8. Quality Assurance Guidelines

8.1 Method detection limits have been established per US EPA Appendix B to Part 136-Definition and Procedure for the Determination of the Method Detection Limit-Revision 1.11. 40 CFR Part 136, 43430, *Federal Register*, Volume 49, No.209. Method detection limits and acceptable linear ranges are presented in Table 1.

8.2 Duplicate injections are performed 10% of the time.

8.3 Duplicate sample analyses are performed 10% of the time if extra sampling can be scheduled.

9. Precision and Accuracy

There is no data available at this time for precision and accuracy.

10. References

This procedure is similar to that published by Supelco, Inc., Analyzing Mixtures of Permanent Gases and Light (C₁-C₄) Hydrocarbons on a Single GC Column, GC Bulletin 712F, Supelco, Inc., Bellefonte, PA, 1983.

Table 1. Method Detection Limits (MDLs), Linear Ranges (% V/V) and Linear Correlation Coefficients for Permanent Gases Analyzed using Porapak QS and TCD detection (11-9-89).

<u>Component</u>	<u>MDL</u>	<u>Linear Range</u>	<u>Corr. Coef.</u>
Carbon Dioxide	0.2 %	0%-40 %	0.9998
Methane	0.02%	0%-100%	0.9995
Air	0.2 %	-----	-----

600C: GAS CHROMATOGRAPHIC ANALYSIS OF ACETONITRILE IN AIR AND GAS SAMPLES USING A PACKED COLUMN

INTRODUCTION

Acetonitrile is analyzed by packed column gas chromatography with a flame ionization detector (FID). The detector signal is collected by a data system for identification and quantitation of acetonitrile.

Procedure 600C: Analysis of Acetonitrile in Air And Gas Samples (Jan 2,1992)

1. Scope and Application
 - 1.1 This instrumental procedure is to be used for the identification of acetonitrile.
 - 1.2 Ambient air, flare exhaust, landfill surface gas, landfill boundary probe, digester gas, landfill gas, engine exhaust and other sample types are analyzed according to this procedure.
2. Summary of Procedure
 - 2.1 The sample is injected, cryogenically trapped and thermally desorbed onto a cooled gas chromatographic column.
 - 2.2 The gas chromatograph is temperature programmed to separate the components. The FID detector is used to analyze acetonitrile.
3. Sample Handling and Preservation
 - 3.1 Tedlar (polyvinylfluoride) sample bags are bought commercially. The bags are filled with ultrapure nitrogen, allowed to sit overnight, flushed and filled again. This process is repeated ten or more times and then a sample of the blank nitrogen in the bag is analyzed after remaining in the bag overnight. At least two bags from every lot of ten are randomly selected to represent the lot. If the two bags are clean, then the entire set is placed in service. The new bag data is compared to previous lots of new bags to determine acceptability. If the bags are contaminated, they are returned for another round of flushing with ultrapure nitrogen and tested again for suitability.
 - 3.2 Tedlar sample bags are leak checked after each use. Each sample bag is filled with ultrapure nitrogen and allowed to sit overnight with a weight on it. If the bag's volume is reduced significantly, the bag is either repaired or discarded.
 - 3.3 After leak checking, the Tedlar sample bags are cleaned. Each bag is flushed and filled three times with ultrapure nitrogen and allowed to sit overnight. This is repeated two more times and then the bag is returned to service.

- 3.4 The Tedlar sample bags are dedicated for specific sample type and segregated for storage to avoid contamination. Presently we have four sample bag types:
 - 3.4.1 Ambient air bags for ambient air, landfill surface gas, landfill boundary probes, and flare exhaust samples.
 - 3.4.2 Raw bags for raw digester and landfill gas.
 - 3.4.3 Scrubber bags for sewer gas.
 - 3.4.4 Scrubber exhaust bags for exhaust from sewer gas scrubbers.
 - 3.5 Samples are collected in the field in Tedlar bags and transported to the laboratory in cardboard boxes to minimize photochemical decomposition. Sample bags are flushed with ultrapure nitrogen (either in the laboratory or the field office) just prior to sample collection.
 - 3.6 Samples must be analyzed within 72 hours, but are normally analyzed the day of sampling or the next day.
4. Apparatus
- 4.1 Gas Chromatography
 - 4.1.1 Varian 6000 gas chromatograph equipped with a flame ionization detector connected to column B.
 - 4.1.2 Column: 2 m x 2 mM ID Pyrex packed with 60/80 mesh Carbopack B coated with 1% SP-1000.
 - 4.1.3 Operating parameters:

Reaction gas for FID:
Fuel gas: Hydrogen (Ultra high purity), 25 mL/min.
Oxidant: Hydrocarbon free air, 300 mL/min.
Carrier gas: Helium (carrier grade), 25 mL/min
Sweep gas: Helium (carrier grade), 40 mL/min
Injector temperature: 150 C
Detector base temperature: 300 C for GC 1 and 200 C for GC2
Range 10E-12 for FID
Temperature program for GC 1: A 10-min hold at an initial temperature of -99 C. Program to 40 C at 40 C/min and hold for 5 min. Program to 220 C at 8 C/min and hold for 14 min.
Temperature program for GC 2: Hold 11 min at -20 C. Program to 30 C at 10 C/min and hold for 12 min then program to 65 C at 8 C/min and hold for 12 min. Program to 170 C at 20 C/min and hold for 5 min. Program to 220 C at 40 C/min and hold until all components have eluted.
 - 4.2 Data System:
 - 4.2.1 Nelson Analytical (3000 series) chromatography software system based on an

IBM AT or PS/2 personal computer with a two channel capacity buffer box (760 series) to collect the raw data from the FID. The data system draws baselines and identifies components according to the software and preselected parameters. The data may be recalculated by the operator by changing the drawn baselines.

- 4.2.2 NWA Quality Analyst version 3.1 statistical program for quality control charting and analysis.
 - 4.2.3 Analytical Automation Specialists AUTOSQC software package which integrates Nelson Analytical 3000 software with NWA Quality Analyst to provide Shewhart quality control charts from chromatographic data.
 - 4.2.4 "Limitgc1" and "SP1L20" are basic programs written using the BASIC interpreter. "Limitgc1" is designed to set up tables with component names and calculated acceptable area count range for these components and then produce a hard copy of these tables. "SP1L20.bas" is a basic program that is used to check the daily 100 mL standard injection. Lines 10-330 were modified from Nelson Analytical's "User.Bas" program. "SP1L20.Bas" imports ASCII data from previously saved area tables, compares the proper areas with the limits set up by "Limitgc1.Bas", and prints an exception report listing all compounds that exceed the limit.
- 4.3 Strip Chart Recorder: Linear Instruments Model 282/MM with dual pens.
 - 4.4 Cryotrap System
 - 4.4.1 Plumbing: 1/8" nickel tubing kept at approximately 200 using 3/16" nichrome wire heating cord controlled by a Glas-Col Model PL-312 Minitrol voltage regulator set at 2.
 - 4.5 Sample Loading System
 - 4.5.1 A 100 mL glass syringe with a Teflon plunger is used to measure the larger sample volumes. Volumes less than 100 mL are loaded onto the cryotrap by injecting the volume through a septum with a gas tight syringe.
 - 4.5.2 Trap utilizing 1/4" ID nickel tubing in a "U" configuration packed with silanized glass beads. The trap is wrapped with 3/16" nichrome wire heating cord controlled by a Glas-Col Model PL-312 Minitrol temperature controller. The trap can be heated to approximately 400 C. Trap B is used for the column connected to the FID (side B).

5. Reagents

- 5.1 Standard mixture was obtained from Scott-Marrin, Inc. in a cylinder at nominal concentrations of 100 ppb V/V in nitrogen. Table I shows the components in the

Scott-Marrin standard currently in use and their concentrations. Working standard is prepared by transferring a volume of 100 ppb V/V standard to a Tedlar bag. See Table II for components and concentrations.

6. Procedure

- 6.1 The data collection method is loaded into the interface box prior to the start of the run. The data collection method instructs the data system how to acquire the data to be generated during the run (i.e. during what times and at what rate the data should be collected). The data system controls all valves through timed event relays.
- 6.2 The cryotrap is immersed into a liquid nitrogen bath for 1 min prior to sample trapping. The valve actuator is placed in the load position. This allows the helium sweep gas to flow through the cryotrap systems and maintains helium carrier flow through the column.
- 6.3 A gaseous sample of up to 500 mL is loaded onto the cryotrap side B with a gas tight syringe. The sweep gas carries sample onto the cryotrap.
- 6.4 The gas chromatograph oven should be cooled to -99 C and ready for injection.
- 6.5 Three min after the gas sample has been loaded onto the cryotrap, the liquid nitrogen bath is removed. After waiting 15 sec (to allow the CO₂ and NO_x to vent) the run is started.
- 6.6 When the run is started several electrically controlled events occur. The valve controlling the sample flow is switched to "inject", allowing helium sweep gas to carry the sample from the trap to the gas chromatograph column. The cryotrap is heated to approximately 400 C for 4 min and then allowed to cool for 3 min while sweeping sample from the trap onto the column. At the end of this sequence the sample valve is switched back to "load".
- 6.7 Periodically, 10-point calibration curves are generated for the FID detector. The lifetime of this curve is determined by detector performance. Retention time is updated on a daily basis.

7. Calculations

- 7.1 The data system identifies and quantifies sample components.
- 7.2 Ten-point calibration curve (point-to-point) is established for acetonitrile. A correction factor may be applied if the daily response is found to vary by more than 20% from when the calibration curve was established.

7.3 The peaks that are identified are calculated by the data system as follows:

$$\text{Conc. of sample in ppb, V/V} = \frac{\text{sample peak area}}{\text{sample Vol, mL}} \times \frac{\text{ppb std} \times \text{mLs of std}}{\text{std area}}$$

x correction factor (if needed)

$$\text{correction factor} = \frac{\text{calibration curve area count}}{\text{daily standard area count}}$$

8. Quality Assurance

- 8.1 The acceptability of the calibration curve is monitored by injecting 100 mL of the Scott-Marrin gas standard each day the gas chromatograph is used. The area count from the standard is entered into the AUTOSQC statistical program. If the daily standard does not meet the statistical criteria, then corrective action is taken.
- 8.2 Weekly QA includes trip blanks (Tedlar bags filled with ultrapure nitrogen that travel with sample bags to the sampling destination and back to the laboratory for analysis), duplicate samples from the sampling site, and duplicate injections of the same sample (to validate laboratory precision).
- 8.3 Method detection limits have not been documented. Presently a "quantitation limit" is used where a value is not reported if the signal is lower than the lowest signal used in the calibration curve. Quantitation limits are raised when high blank values are found. Typical values are listed in Table I, for both landfill gas (0.5 mL) and ambient air (500 mL) injection volumes.
- 8.4 Sample spiking techniques have not been established.

Table I - Component Currently Analyzed Using FID Cryotrap Method, Test Code and the Current Limits of Quantitation in ppb V/V.

		<u>Limit of Quantitation</u>		
<u>Test Code</u>	<u>Component</u>	<u>sample volume 0.5 mL</u>	<u>sample volume 500 mL</u>	<u>Detector Used</u>
665	Acetonitrile	5000	5.0	FID

Table II - Scott-Marrin Standard Currently Used And Its Concentrations.

<u>Component</u>	<u>ppb V/V</u>
Acetonitrile	90
Nitrogen	BALANCE

600B: GAS CHROMATOGRAPHIC ANALYSIS OF ORGANIC COMPOUNDS IN AIR AND GAS SAMPLES USING A PACKED COLUMN

INTRODUCTION

Volatile organic compounds are analyzed by packed column gas chromatography with a photoionization detector (PID) in series with a Hall electrolytic conductivity detector (HECD). The detector signals are collected by a data system for identification and quantitation of the sample components.

Procedure 600B: Analysis of Organic Compounds in Air and Gas (Jan 2, 1992)

1. Scope and Application
 - 1.1 This instrumental procedure is to be used for the identification and quantitation of volatile organic compounds. The components currently analyzed by this method and their test codes are listed in Table I.
 - 1.2 Ambient air, flare exhaust, landfill surface gas, landfill boundary probe, digester gas, landfill gas, engine exhausts and other sample types are analyzed according to this procedure.
2. Summary of Procedure
 - 2.1 The sample is injected, cryogenically trapped and thermally desorbed onto a cooled gas chromatographic column.
 - 2.2 The gas chromatograph is temperature programmed to separate the components. Aromatics and unsaturated hydrocarbons are analyzed using the PID. The effluent from the PID is analyzed for halocarbons by HECD.
3. Sample Handling and Preservation
 - 3.1 Tedlar (polyvinylfluoride) sample bags are bought commercially. The bags are filled with ultrapure nitrogen, allowed to sit overnight, flushed and filled again. This process is repeated ten or more times and then a sample of the blank nitrogen in the bag is analyzed after remaining in the bag overnight. At least two bags from every lot of ten are randomly selected to represent the lot. If the two bags are clean, then the entire set is placed in service. The new bag data is compared to previous lots of new bags to determine acceptability. If the bags are contaminated, they are returned for another round of flushing with ultrapure nitrogen and tested again for suitability.
 - 3.2 Tedlar sample bags are leak checked after each use. Each sample bag is filled with ultrapure nitrogen and allowed to sit overnight with a weight on it. If the bag's volume is reduced significantly, the bag is either repaired or discarded.
 - 3.3 After leak checking, the Tedlar sample bags are cleaned. Each bag is flushed and

filled three times with ultrapure nitrogen and allowed to sit overnight. This is repeated two more times and then the bag is returned to service.

3.4 The Tedlar sample bags are dedicated for specific sample type and segregated for storage to avoid contamination. Presently we have four sample bag types:

3.4.1 Ambient air bags for ambient air, landfill surface gas, landfill boundary probes, and flare exhaust samples.

3.4.2 Raw bags for raw digester and landfill gas.

3.4.3 Scrubber bags for sewer gas.

3.4.4 Scrubber exhaust bags for exhaust from sewer gas scrubbers.

3.5 Samples are collected in the field in Tedlar bags and transported to the laboratory in cardboard boxes to minimize photochemical decomposition. Sample bags are flushed with ultrapure nitrogen (either in the laboratory or the field office) just prior to sample collection.

3.6 Samples must be analyzed within 72 hours, but are normally analyzed the day of sampling or the next day.

4. Apparatus

4.1 Gas Chromatography

4.1.1 Varian 6000 gas chromatograph equipped with a 10.2 eV HNu model PI-52-02 photoionization detector in series with a Hall model 700A electrolytic conductivity detector (Tracor) operated in the halogen mode.

4.1.2 Column: 2 m x 2 mM ID Pyrex packed with 60/80 mesh Carbopack B coated with 1% SP-1000.

4.1.3 Operating parameters:

Carrier gas: Helium (carrier grade), 25 mL/min

Sweep gas: Helium (carrier grade), 40 mL/min

Injector temperature: 150 C

Detector base temperature: 300 C

Range: 10 for PID and HECD

HECD electrolyte: n-propanol at 0.5 mL/min

Temperature program: A 10-min hold at an initial temperature of -99 C. Program to 40 C at 40 C/min and hold for 5 min. Program to 220 C at 8 C/min and hold for 14 min.

4.2 Data System:

4.2.1 Nelson Analytical (3000 series) chromatography software system based on an

IBM AT or PS/2 personal computer with a two channel capacity buffer box (760 series) to collect the raw data from both PID and HECD. The data system draws baselines and identifies components according to the software and preselected parameters. The data may be recalculated by the operator by changing the drawn baselines.

- 4.2.2 NWA Quality Analyst version 3.1 statistical program for quality control charting and analysis.
 - 4.2.3 Analytical Automation Specialists AUTOSQC software package which integrates Nelson Analytical 3000 software with NWA Quality Analyst to provide Shewhart quality control charts from chromatographic data.
 - 4.2.4 "Limitgc1" and "SP1L20" are basic programs written using the BASIC interpreter. "Limitgc1" is designed to set up tables with component names and calculated acceptable area count ranges for these components and then produce a hard copy of these tables. "SP1L20.bas" is a basic program that is used to check the daily 50 mL standard injection. Lines 10-330 were modified from Nelson Analytical's "User.Bas" program. "SP1L20.bas" imports ASCII data from previously saved area tables, compares the proper areas with the limits set up by "Limitgc1.bas", and prints an exception report listing all compounds that exceed the limit.
- 4.3 Strip Chart Recorder: Linear Instruments Model 282/MM with dual pens for PID and HECD.
- 4.4 Cryotrap System
- 4.4.1 Plumbing: 1/8" nickel tubing kept at approximately 200 °C using 3/16" nichrome wire heating cord controlled by a Glas-Col Model PL-312 Minitrol voltage regulator set at 2.
 - 4.4.2 1/16" ID Nafion tubing leading from the trap to a stainless steel Valco E60 injection valve. The Nafion tubing (a perfluorinated monomer membrane) is available from Perma Pure Products, Inc., Oceanport, N.J. Water vapor is selectively diffused through the membrane (References 1 and 2).
- 4.5 Sample Loading System
- 4.5.1 A 100 mL glass syringe with a Teflon plunger is used to measure the larger sample volumes. Volumes less than 100 mL are loaded onto the cryotrap by injecting the volume through a septum with a gas tight syringe.
 - 4.5.2 Trap utilizing 1/4" ID nickel tubing in a "U" configuration packed with silanized glass beads. The trap is wrapped with 3/16" nichrome wire heating cord controlled by a Glas-Col Model PL-312 Minitrol temperature controller. The trap can be heated to approximately 400°C.

5. Reagents

- 5.1 A 1,3-butadiene and benzyl chloride standard mixture was obtained in a Scott "Aculife" polished aluminum cylinder from Scott Specialty Gases with nitrogen as a balance gas. Working standards are prepared by transferring a volume of 100 ppb V/V nominal standard to a Tedlar bag. See Table III for concentrations.
- 5.2 Standard mixtures were obtained from Scott-Marrin, Inc. in cylinders at nominal concentrations of 100 ppb V/V in nitrogen. Table II shows the components in the three Scott-Marrin standards currently in use and their concentrations. Working standards are prepared by transferring a volume of 100 ppb V/V standard to a Tedlar bag.
- 5.3 Quality assurance standards obtained from the National Institute of Standards and Technology (NIST) at nominal concentrations of 250 and 5.0 ppb V/V. The components and concentrations of the nominal 250 ppb V/V standards are listed in Table IV and the nominal 5.0 ppb V/V standards are listed in Table V. NIST gas standards are certified for 2 years.

6. Procedure

- 6.1 The data collection methods are loaded into the interface boxes prior to the start of the run. The data collection method instructs the data system how to acquire the data to be generated during the run (i.e. during what times and at what rate the data should be collected). The data system controls all valves through timed event relays.
- 6.2 The cryotrap is immersed into a liquid nitrogen bath for 1 min prior to sample trapping. The valve actuator is placed in the load position. This allows the helium sweep gas to flow through the cryotrap system and maintains helium carrier flow through the column.
- 6.3 A gaseous sample of up to 500 mL is loaded onto the cryotrap with a gas tight syringe. The sweep gas carries sample onto the cryotrap.
- 6.4 The gas chromatograph oven should be cooled to -99 C and ready for injection.
- 6.5 Three min after the gas sample has been loaded onto the cryotrap, the liquid nitrogen bath is removed. After waiting 15 sec (to allow the CO₂ and NO_x to vent) the run is started.
- 6.6 When the run is started, several electrically controlled events occur. The valve controlling the sample flow is switched to "inject", allowing helium sweep gas to carry the sample from the trap to the gas chromatograph column. The cryotrap is heated to approximately 400 C for 4 min and then allowed to cool for 3 min while sweeping sample from the trap onto the column. At the end of this sequence the sample valve is switched back to "load".
- 6.7 Periodically, 10-point calibration curves are generated for each detector. The lifetime

of these curves is determined by detector performance. Retention times are updated on a daily basis.

7. Calculations

7.1 The data system identifies and quantifies sample components.

7.2 Ten-point calibration curves (point-to-point) are established for each detector and component. Correction factors may be applied if the daily response is found to vary by more than 20% from when the calibration curve was established.

7.3 The peaks that are identified are calculated by the data system as follows:

$$\text{Conc. of sample in ppb, V/V} = \frac{\text{sample peak area}}{\text{sample Vol, mLs}} \times \frac{\text{ppb std}}{\text{std area}} \times \frac{\text{mLs of std}}{\text{std area}} \\ \times \text{correction factor (if needed)}$$

$$\text{correction factor} = \frac{\text{calibration curve area count}}{\text{daily standard area count}}$$

8. Quality Assurance

8.1 The acceptability of the calibration curves are monitored by injecting 50 mL of the Scott-Marrin and the Scott Specialty gas standards each day the gas chromatograph is used. The area counts from the standards are entered into the AUTOSQC statistical program. If the daily standards do not meet the statistical criteria, then corrective action is taken.

8.2 Weekly QA includes trip blanks (Tedlar bags filled with ultrapure nitrogen that travel with sample bags to the sampling destination and back to the laboratory for analysis), duplicate samples from the sampling site, and duplicate injections of the same sample (to validate laboratory precision).

8.3 Method detection limits have not been documented. Presently a "quantitation limit" is used where a value is not reported if the signal is lower than the lowest signal used in the calibration curve. Quantitation limits are raised when high blank values are found. Typical values are listed in Table I, for both landfill gas (0.5 mL) and ambient air (500 mL) injection volumes.

8.4 Sample spiking techniques have recently been established. Spiked samples are run weekly and the data is entered into a data base.

8.5 Mixtures obtained from the National Institute of Standards and Technology are run weekly. The results are entered into the AUTOSQC statistical program. If the results do not meet statistical criteria, then corrective action is taken.

Table II - Scott-Marrin Standards Currently Used And Their Concentrations.

Scott-Marrin Standard I

<u>Component</u>	<u>ppb V/V</u>
Vinyl Chloride	99.5
Carbon Tetrachloride	103
1,1-Dichloroethylene	106
cis-1,2-Dichloroethylene	109
Tetrachloroethylene	100
Benzene	103
Toluene	115
1,2-Dibromoethane	111
Acetonitrile	90
Nitrogen	BALANCE

Scott-Marrin Standard II

<u>Component</u>	<u>ppb V/V</u>
Methylene Chloride	99.5
Chloroform	106
1,1-Dichloroethane	104
1,2-Dichloroethane	102
1,1,1-Trichloroethane	103
Trichloroethylene	100
m+p-Xylene	214
Chlorobenzene	103
m-Dichlorobenzene	115
o-Dichlorobenzene	115
o-Xylene	106
Nitrogen	BALANCE

Scott-Marrin Standard III

<u>Component</u>	<u>ppb V/V</u>
Freon-11	100.3
Nitrogen	BALANCE

Table III - Scott Specialty Gases Standard

<u>Component</u>	<u>ppb V/V</u>
1,3-Butadiene	94.7
Benzyl Chloride	109
Nitrogen	BALANCE

Table IV - High Level NIST Standards Currently Used And Their Concentrations

Halocarbon Standard

<u>Component</u>	<u>ppb V/V</u>
Carbon Tetrachloride	245
Chloroform	247
Tetrachloroethylene	250
Vinyl Chloride	241
Nitrogen	BALANCE

Aromatic Standard

<u>Component</u>	<u>ppb V/V</u>
Benzene	238
Toluene	225
Chlorobenzene	223
Bromobenzene	226
Nitrogen	BALANCE

Table V - Low Level NIST Standards Currently Used And Their Concentrations

<u>Component</u>	<u>ppb V/V</u>
Vinyl Chloride	5.3
Dichloromethane	5.0
Chloroform	5.0
1,2-Dichloroethane	5.0
1,1,1-Trichloroethane	5.1
Benzene	5.0
Carbon Tetrachloride	5.0
Trichloroethylene	5.0
Toluene	4.9
Tetrachloroethylene	5.0
Chlorobenzene	5.0
O-Xylene	5.0
1,3-Butadiene	4.7 (not certified)
1,2-Dibromoethane	4.8 (not certified)
Nitrogen	BALANCE

APPENDIX B
QUALITY ASSURANCE

Appendix B.1
Quality Assurance Program Summary

QUALITY ASSURANCE PROGRAM SUMMARY AND ARB CERTIFICATION

Carnot ensures the quality and validity of its emission measurement and reporting procedures through a rigorous quality assurance (QA) program. The program is developed and administered by an internal QA Officer and encompasses seven major areas:

1. Development and use of an internal QA manual.
2. QA reviews of reports, laboratory work, and field testing.
3. Equipment calibration and maintenance.
4. Chain of custody.
5. Training.
6. Knowledge of current test methods.
7. Agency certification.

Each of these areas is discussed individually below.

Quality Assurance Manual. Carnot has prepared a QA Manual according to EPA guidelines. The manual serves to document and formalize all of Carnot's QA efforts. The manual is constantly updated, and each member of the Source Test Division is required to read and understand its contents. The manual includes details on the other six QA areas discussed below.

QA Reviews. Carnot's review procedure includes review of each source test report by the QA Officer, and spot check reviews of laboratory and field work.

The most important review is the one that takes place before a test program begins. The QA Officer works closely with Source Test Division personnel to prepare and review test protocols. Test protocol review includes selection of appropriate test procedures, evaluation of any interferences or other restrictions that might preclude use of standard test procedures, and evaluation and/or development of alternate procedures.

Equipment Calibration and Maintenance. The equipment used to conduct the emissions measurements is maintained according to the manufacturer's instructions to ensure proper operation. In addition to the maintenance program, calibrations are carried out on each measurement device according to the schedule outlined by the California Air Resources Board (CARB). The schedule for maintenance and calibrations are given in Tables B-1 and B-2. Quality control checks are also conducted in the field for each test program. The following is a partial list of checks made as part of each CEM system test series.

- Sample acquisition and conditioning system leak check.
- 2-point analyzer calibrations (all analyzers)
- 3-point analyzer calibrations (analyzers with potential for linearity errors).
- Complete system calibration check ("dynamic calibration" through entire sample system).

- Periodic analyzer calibration checks (once per hour) are conducted at the start and end of each test run. Any change between pre- and post-test readings are recorded.
- All calibrations are conducted using gases certified by the manufacturer to be + 1% of label value (NBS traceable).

Calibration and CEM performance data are fully documented, and are included in each source test report.

Chain of Custody. Carnot maintains full chain of custody documentation on all samples and data sheets. In addition to normal documentation of changes between field sample custodians, laboratory personnel, and field test personnel, Carnot documents every individual who handles any test component in the field (e.g., probe wash, impinger loading and recovery, filter loading and recovery, etc.).

Samples are stored in a locked area to which only Source Test Division personnel have access. Neither other Carnot employees nor cleaning crews have keys to this area.

Data sheets are copied immediately upon return from the field, and this first generation copy is placed in locked storage. Any notes made on original sheets are initialed and dated.

Training. Personnel training is essential to ensure quality testing. Carnot has formal and informal training programs which include:

1. Attendance at EPA-sponsored training courses.
2. Enrollment in EPA correspondence courses.
3. A requirement for all technicians to read and understand Carnot's QA Manual.
4. In-house training and QA meetings on a regular basis.
5. Maintenance of training records.

Knowledge of Current Test Methods. With the constant updating of standard test methods and the wide variety of emerging test methods, it is essential that any qualified source tester keep abreast of new developments. Carnot subscribes to services which provide updates on EPA and CARB reference methods, and on EPA, CARB and SCAQMD rules and regulations. Additionally, source test personnel regularly attend and present papers at testing and emission-related seminars and conferences. Carnot personnel maintain membership in the Air Pollution Control Association, the Source Evaluation Society, and the ASME Environmental Control Division.

AGENCY CERTIFICATION

Carnot is certified by the CARB as an independent source test contractor for gaseous and particulate measurements. Carnot also participates in EPA QA audit programs for Methods 5, 6 and 7.

TABLE B-1
 SAMPLING INSTRUMENTS AND EQUIPMENT CALIBRATION SCHEDULE
 As Specified by the CARB

Instrument Type	Frequency of Calibration	Standard of Comparison or Method of Calibration	Acceptance Limits
Orifice Meter (large)	12 months	Calibrated dry test meter	± 2% of volume measured
Dry Gas Meter	12 months or when repaired	Calibrated dry test meter	± 2% of volume measured
S-Type Pitot (for use with EPA-type sampling train)	6 months	EPA Method 2	Cp constant (+5%) over working range; difference between average Cp for each leg must be less than 2%
Vacuum Gauges Pressure Gauges	6 months	Manometer	± 3%
Field Barometer	6 months	Mercury barometer	± 0.2" Hg
Temperature Measurement	6 months	NBS mercury thermometer or NBS calibrated platinum RTD	± 4°F for <400°F ± 1.5% for >400°F
Temperature Readout Devices	6 months	Precision potentiometer	± 2% full scale reading
Analytical Balance	12 months (check prior to each use)	Should be performed by manufacturer or qualified laboratory	± 0.3 mg of stated weight
Probe Nozzles	12 Months	Nozzle diameter check micrometer	Range < ± 0.10 mm for three measurements
Continuous Analyzers	Depends upon use, frequency and performance	As specified by manufacturers operating manuals, EPA NBS gases and/or reference methods	Satisfy all limits specified in operating specifications

TABLE B-2
EQUIPMENT MAINTENANCE SCHEDULE
Based on Manufacturer's Specifications and Carnot Experience

Equipment	Performance Requirement	Maintenance Interval	Corrective Action
Pumps	<ol style="list-style-type: none"> 1. Absence of leaks 2. Ability to draw manufacturer required vacuum and flow 	Every 500 hours of operation or 6 months, whichever is less	<ol style="list-style-type: none"> 1. Visual inspection 2. Clean 3. Replace worn parts 4. Leak check
Flow Measuring Device	<ol style="list-style-type: none"> 1. Free mechanical movement 2. Absence of malfunction 	Every 500 hours of operation or 6 months, whichever is less After each test, if used in H ₂ S sampling or other corrosive atmospheres	<ol style="list-style-type: none"> 1. Visual inspection 2. Clean 3. Calibrate
Sampling Instruments	<ol style="list-style-type: none"> 1. Absence of malfunction 2. Proper response to zero, span gas 	As required by the manufacturer	As recommended by manufacturer
Integrated Sampling Tanks	Absence of leaks	Depends on nature of use	<ol style="list-style-type: none"> 1. Steam clean 2. Leak check
Mobile Van Sampling Systems	Absence of leaks	Depends on nature of use	<ol style="list-style-type: none"> 1. Change filters 2. Change gas dryer 3. Leak check 4. Check for system contamination
Sampling Lines	Sample degradation less than 2%	After each test or test series	Blow filtered air through line until dry

Appendix B.2
ARB Certification

State of California
AIR RESOURCES BOARD

Executive Order G-767

Approval to Carnot
To Conduct Testing as an Independent Contractor

WHEREAS, the Air Resources Board ("Board"), pursuant to Section 41512 of the California Health and Safety Code, has established the procedures contained in Section 91200-91220, Title 17, California Code of Regulations, to allow the use of independent testers for compliance tests required by the Board; and


WHEREAS, pursuant to Sections 91200-91220, Title 17, California Code of Regulations, the Executive Officer has determined that Carnot meets the requirements of the Board for conducting ARB Test Methods 1, 2, 3, 4, 5, 6, 8, 10, and 100 (NOx, O2) when the following conditions are met:

1. Carnot conducts ARB Test Method 100 for O2 using a Teledyne 326 analyzer with either a A5 or a B1 sensor, or a paramagnetic analyzer.

NOW, THEREFORE, BE IT ORDERED that Carnot is granted an approval, from the date of execution of this order, until June 30, 1993 to conduct the tests listed above, subject to compliance with Section 91200-91220, Title 17, California Code of Regulations.

BE IT FURTHER ORDERED that during the approved period the Executive Officer or his or her authorized representative may field audit one or more tests conducted pursuant to this order for each type of testing listed above.

Executed this 10TH day of August 1992, at Sacramento, California.



James J. Morgester, Chief
Compliance Division


State of California
Air Resources Board
Approved Independent Contractor

Carnot


This is to certify that the company listed above has been approved by the Air Resources Board to conduct compliance testing, pursuant to Section 91207, Title 17, California Code of Regulations, until June 30, 1993, for those test methods listed below:

ARB Source Test Methods:

1, 2, 3, 4, 5, 6, 8, 10, 100 (NO_x, O₂),



James J. Morgester, Chief
Compliance Division



for H. Royce Jackson, Manager
Field Evaluation Section

Appendix B.3
Calibration Data

CARNOT

DRY GAS METER CALIBRATION

FIELD GAS METER I.D....ES-28

TEST METER ID..... ES-13

DATE.....12/22/92

TEST METER LAST CAL..12/08/92

BAROMETRIC PRESSURE....30.18

TEST METER CAL FACTOR 1.0006 = Yt

CALIBRATION BY..... BC

LEAK CHECK.....GOOD @ 8 IN. H2O VAC

FIELD METER				TEST METER				RESULTS			
VOLUME cu.ft.	TEMP.a 'F	DELTA H "H2O	TIME min.	VOLUME cu.ft.	TEMP. 'F	PRESS. "H2O	Q cfm	Y	H@	AVE.Y	AVG. H@
6.062	61.8	0.50	15	5.962	54.5	-0.17	0.404	0.996	1.70	1.0002	1.72
6.062	64.5	0.50	15	5.957	54.5	-0.17	0.404	1.001	1.69		
5.952	67.5	0.50	15	5.843	55.5	-0.17	0.397	1.004	1.75		
5.941	71.3	0.80	12	5.823	56.0	-0.21	0.495	1.007	1.80	1.0066	1.80
5.955	72.0	0.80	12	5.826	56.5	-0.21	0.496	1.006	1.80		
5.985	72.3	0.80	12	5.858	56.5	-0.21	0.499	1.007	1.78		
5.536	74.0	1.00	10	5.413	57.5	-0.25	0.554	1.006	1.81	1.0074	1.82
5.536	75.8	1.00	10	5.414	57.5	-0.25	0.554	1.010	1.81		
5.531	74.8	1.00	10	5.402	58.0	-0.25	0.553	1.006	1.82		
5.322	76.3	2.00	7	5.296	58.0	-0.46	0.760	1.025	1.86	1.0130	1.92
5.317	75.5	2.00	7	5.210	58.0	-0.46	0.760	1.008	1.93		
5.260	66.8	2.00	7	5.266	61.0	-0.46	0.751	1.007	1.94		
5.545	74.5	3.00	6	5.539	62.0	-0.63	0.924	1.014	1.92	1.0111	1.92
5.585	74.3	3.00	6	5.550	62.0	-0.63	0.931	1.009	1.92		
5.623	76.8	3.00	6	5.570	62.0	-0.63	0.937	1.010	1.89		

AVERAGE = 1.0076 1.83
 AVERAGE cfm @H=1.0 = 0.553

CARNOT

DRY GAS METER CALIBRATION

FIELD GAS METER I.D....ES-42

TEST METER ID..... ES-13

DATE.....12/28/92

TEST METER LAST CAL..12/08/92

BAROMETRIC PRESSURE....29.91

TEST METER CAL FACTOR 1.0006 = Yt

CALIBRATION BY..... BC

LEAK CHECK.....GOOD @ 8 IN. H2O VAC

FIELD METER				TEST METER				RESULTS			
VOLUME cu. ft.	TEMP. a 'F	DELTA H "H2O	TIME min.	VOLUME cu. ft.	TEMP. 'F	PRESS. "H2O	Q cfm	Y	H@	AVE. Y	AVG. H@
5.869	67.8	0.50	15	5.776	68.0	-0.13	0.391	0.983	1.90	0.9899	1.88
5.886	74.5	0.50	15	5.761	65.3	-0.13	0.392	0.995	1.86		
5.926	75.5	0.50	15	5.780	66.0	-0.13	0.395	0.992	1.85		
5.911	76.8	0.80	12	5.791	66.5	-0.20	0.493	0.997	1.89	0.9960	1.93
5.858	78.0	0.80	12	5.711	66.5	-0.20	0.488	0.994	1.94		
5.861	78.3	0.80	12	5.725	66.5	-0.20	0.488	0.997	1.93		
5.781	79.0	1.00	10	5.653	66.5	-0.25	0.578	0.999	1.72	0.9938	1.83
5.381	79.3	1.00	10	5.235	67.0	-0.25	0.538	0.993	2.01		
5.799	79.0	1.00	10	5.625	67.0	-0.25	0.580	0.990	1.74		
5.258	78.8	2.00	7	5.141	66.0	-0.40	0.751	0.996	2.05	0.9969	2.00
5.316	80.0	2.00	7	5.182	66.0	-0.40	0.759	0.995	2.01		
5.408	79.8	2.00	7	5.294	66.0	-0.40	0.773	0.999	1.93		
5.501	80.5	3.00	6	5.400	66.0	-0.58	0.917	1.000	2.05	1.0007	2.03
5.558	80.8	3.00	6	5.455	66.0	-0.58	0.926	1.001	2.01		
5.546	81.8	3.00	6	5.433	66.0	-0.58	0.924	1.001	2.02		

AVERAGE = 0.9954 1.93
 AVERAGE cfm @H=1.0 = 0.565

CARNOT SPAN GAS RECORD

CLIENT/LOCATION: PVER6 DATE: 12-30-92
 BY: CHF

GAS	SPAN CYLINDER		AUX. SPAN CYLINDER	
	CYLINDER NO.	CONCENTRATION	CYLINDER NO.	CONCENTRATION
ZERO	AAL 1254	99.998 N ₂		
NOx	ALM006719	12.49 NO 12.68 NOx	AAL 18584	22.43 NO 22.81 NOx
O ₂	ALM006709	5.162	AAL 3966	8.013
CO	ALM015919	11.22	A 1445	16.17
CO ₂	ALM006709	15.15	AAL 3966	22.74
SO ₂				

CARNOT INSTRUMENT LINEARITY

	ANALYZER				
	O ₂	CO ₂	CO	NOx	SO₂ CO
ANALYZER RANGE	0-10	0-25	0-20	0-25	0-100
SET TO HIGH STD (80-90% OF RANGE)	8.013	22.74/20.00	16.17	22.81	16.17
ACTUAL VALUE OF LOW STD	5.162	15.15/16.00	11.22	12.68	11.22
AS-FOUND LOW STD (50-60% OF RANGE)	5.22	15.35	11.34	12.50	11.4
DIFFERENCE IN % OF FULL SCALE	0.6	0.8	0.6	-0.7	0.2

% ERROR CALCULATION:

$$\frac{(\text{AS FOUND} - \text{ACTUAL VALUE OF SPAN})}{\text{RANGE}} \times 100$$

ALLOWABLE DEVIATION IS 2% OF FULL SCALE (2 SQUARES ON STRIP CHART).



Scott Specialty Gases, Inc.

2600 CAJON BLVD., SAN BERNARDINO, CA 92411

CARNOT
ATTN: RICK MADRIGAL
15991 RED HILL AVE
SUITE 110
TUSTIN CA 92680

FAX 714-887-0549
PHONE: 714-887-2571

Shipped From: Scott SAN BERNARDINO, CA
Date Shipped 11/19/92
Our Project No: 22392
Your P.O. No: 9059
Page 1 of 1
Expiration Date: 5/94

CERTIFICATE OF ANALYSIS - EPA PROTOCOL GASES

Certified Per Traceability Protocol No. 1 Procedure No. G1 Cylinder No. AA118584 Cylinder Pressure 2000 PSIG Certified Accuracy ±1% % NIST Traceable

REFERENCE STD

COMPONENTS	CERTIFIED CONC	SRM/CRM NO.	CYL. NO.	CONC.
NITRIC OXIDE	22.43 PPM	GMIS TRACEABLE TO CRM1684B	ALM017542	48.20 PPM
NOX	22.81 PPM			

GAS ANALYZER

MAKE/MODEL/SERIAL NO.	LAST CAL. DATE	ANALYTICAL PRINCIPLE
THERMO-ELECTRON 10AR 14853-150	10/29/92	CHEMI-LUMINESCENT

BALANCE GAS NITROGEN
ANALYZER READINGS: Z = Zero Gas T = Test Gas R = Reference Gas

Component <u>NITRIC OXIDE</u>			
First Analysis Date	Units	mv	
11/10/92			
Z 0.00	R 58.12	T 27.06	
R 58.10	Z 0.00	T 27.03	
Z 0.06	T 27.02	R 58.02	
		Mean Test Assay	22.43 PPM
Second Analysis Date	Units	mv	
11/17/92			
Z 0.00	R 60.02	T 27.81	
R 60.23	Z -0.08	T 28.08	
Z -0.01	T 27.94	R 60.03	
		Mean Test Assay	22.43 PPM

Component _____			
Date	Units	mv	
Z _____	R _____	T _____	
R _____	Z _____	T _____	
Z _____	T _____	R _____	
		Mean Test Assay	_____
Date	Units	mv	
Z _____	R _____	T _____	
R _____	Z _____	T _____	
Z _____	T _____	R _____	
		Mean Test Assay	_____

Component _____			
Date	Units	mv	
Z _____	R _____	T _____	
R _____	Z _____	T _____	
Z _____	T _____	R _____	
		Mean Test Assay	_____
Date	Units	mv	
Z _____	R _____	T _____	
R _____	Z _____	T _____	
Z _____	T _____	R _____	
		Mean Test Assay	_____

Chronology: Date _____
Assay _____

Analyst G. KING

Approved By: A.F. Lange
A.F. LANGE, Ph.D., MANAGER, QA



Scott Specialty Gases, Inc.

FAX 714-887-0549
PHONE: 714-887-2571

2600 CAJON BLVD., SAN BERNARDINO, CA 92411

CARNOT
ATTN: JIM MULLIGAN
15991 RED HILL AVE
SUITE 110
TUSTIN CA 92680

Shipped From: Scott SAN BERNARDINO, CA
Date Shipped 04/06/92
Our Project No: 18768
Your P.O. No: 8195
Page 1 of 1
Expiration Date: 10/93

CERTIFICATE OF ANALYSIS - EPA PROTOCOL GASES

Certified Per Traceability Protocol No. 1 Procedure No. G1 Cylinder No. ALM006719 Cylinder Pressure 2000 PSIG Certified Accuracy ±1% % NIST Traceable

REFERENCE STD

GAS ANALYZER

COMPONENTS	CERTIFIED CONC	SRM/CRM NO.	CYL. NO.	CONC.	MAKE/MODEL/SERIAL NO.	LAST CAL DATE	ANALYTICAL PRINCIPLE
<u>NITRIC OXIDE</u>	<u>12.49 PPM</u>	<u>SRM2629A</u>	<u>FF28500</u>	<u>19.4 PPM</u>	<u>THERMO-ELECTRON</u>	<u>12/18/91</u>	<u>CHEMI-LUMINESCENT</u>
<u>NOX</u>	<u>12.68 PPM</u>				<u>10AR 14853-150</u>		

BALANCE GAS NITROGEN
ANALYZER READINGS: Z = Zero Gas T = Test Gas R = Reference Gas

Component NITRIC OXIDE

First Analysis Date	<u>03/26/92</u>	Units	<u>mv</u>
Z	<u>0.00</u>	R	<u>71.14</u>
R	<u>71.20</u>	Z	<u>0.74</u>
Z	<u>-0.49</u>	T	<u>45.82</u>
		R	<u>71.29</u>
		Mean Test Assay	<u>12.51 PPM</u>
Second Analysis Date	<u>04/02/92</u>	Units	<u>mv</u>
Z	<u>0.00</u>	R	<u>75.54</u>
R	<u>75.40</u>	Z	<u>-1.25</u>
Z	<u>-1.20</u>	T	<u>48.10</u>
		R	<u>75.36</u>
		Mean Test Assay	<u>12.46 PPM</u>

Component _____

Date	_____	Units	<u>mv</u>
Z	_____	R	_____
R	_____	Z	_____
Z	_____	T	_____
		R	_____
		Mean Test Assay	_____
Date	_____	Units	<u>mv</u>
Z	_____	R	_____
R	_____	Z	_____
Z	_____	T	_____
		R	_____
		Mean Test Assay	_____

Component _____

Date	_____	Units	<u>mv</u>
Z	_____	R	_____
R	_____	Z	_____
Z	_____	T	_____
		R	_____
		Mean Test Assay	_____
Date	_____	Units	<u>mv</u>
Z	_____	R	_____
R	_____	Z	_____
Z	_____	T	_____
		R	_____
		Mean Test Assay	_____

Chronology: Date _____
Assay _____

Analyst G. KING

Approved By: A.F. LANGE
A.F. LANGE, Ph.D., MANAGER, QA



Scott Specialty Gases, Inc.

FAX 714-887-0549
PHONE: 714-887-2571

2600 CAJON BLVD., SAN BERNARDINO, CA 92411

CARNOT
JIM MULLIGAN
15991 RED HILL AVE
SUITE 110
TUSTIN CA 92680

Shipped From: Scott SAN BERNARDINO, CA
Date Shipped 01/24/92
Our Project No: 17560
Your P.O. No: 7967
Page 1 of 1
Expiration Date: 07/93

CERTIFICATE OF ANALYSIS - EPA PROTOCOL GASES

Certified Per Traceability Protocol No. 1 Procedure No. G1 Cylinder No. ALM015919 Cylinder Pressure 2000 PSIG Certified Accuracy +1% % NIST Traceable

REFERENCE STD

GAS ANALYZER

COMPONENTS	CERTIFIED CONC	SRM/CRM NO.	CYL. NO.	CONC.	MAKE/MODEL/SERIAL NO.	LAST CAL DATE	ANALYTICAL PRINCIPLE
CARBON MONOXIDE	11.22 PPM	SRM1677	CAL5078	9.90 PPM	HORIBA AIA-24 4564403	10/23/91	INFRA-RED

BALANCE GAS NITROGEN
ANALYZER READINGS: Z = Zero Gas T = Test Gas R = Reference Gas

Component CARBON MONOXIDE

First Analysis Date	Units	mv
Z <u>00.00</u>	R <u>19.6</u>	T <u>22.2</u>
R <u>19.6</u>	Z <u>00.06</u>	T <u>22.2</u>
Z <u>-00.02</u>	T <u>22.2</u>	R <u>19.6</u>
Mean Test Assay <u>11.25 PPM</u>		
Second Analysis Date	Units	mv
Z <u>00.00</u>	R <u>19.6</u>	T <u>22.1</u>
R <u>19.6</u>	Z <u>00.06</u>	T <u>22.1</u>
Z <u>00.07</u>	T <u>22.1</u>	R <u>19.6</u>
Mean Test Assay <u>11.19 PPM</u>		

Component _____

Date	Units	mv
Z _____	R _____	T _____
R _____	Z _____	T _____
Z _____	T _____	R _____
Mean Test Assay _____		
Date	Units	mv
Z _____	R _____	T _____
R _____	Z _____	T _____
Z _____	T _____	R _____
Mean Test Assay _____		

Component _____

Date	Units	mv
Z _____	R _____	T _____
R _____	Z _____	T _____
Z _____	T _____	R _____
Mean Test Assay _____		
Date	Units	mv
Z _____	R _____	T _____
R _____	Z _____	T _____
Z _____	T _____	R _____
Mean Test Assay _____		

Chronology: Date _____
Assay _____

Analyst M. JOHNSON

Approved By: Jim Ross
JIM ROSS, PLANT MANAGER

114 MULLIGAN
15991 RED HILL AVE
SUITE 110

TUSTIN CA 92680



Scott Specialty Gases, Inc.

FAX: 714-887-0549
PHONE: 714-887-2571

2600 CAJON BLVD., SAN BERNARDINO, CA 92405

Shipped From: Scott N B. RD1, CA
Date Shipped: 05/18/92
Our Project No: 19562
Your P.O. No: 8376
Page 1 of 1
Expiration Date: 11/93

CERTIFICATE OF ANALYSIS - EPA PROTOCOL GASES

Certified Per Traceability Protocol No. 1 Procedure No. G1 Cylinder No. A1445 Cylinder Pressure 2015 PSIG Certified Accuracy ±1% NIST Traceable

REFERENCE STD

GAS ANALYZER

COMPONENTS	CERTIFIED CONC	SPM/CPM NO.	CYL. NO.	CONC.	MAKE/MODEL/SERIAL NO.	LAST CAL DATE	ANALYTICAL PRINCIPLE
CARBON MONOXIDE	16.17 PPM	CRM2635	AA19760	24.30 PPM	HORIBA AIA-24 4564403	02/10/92	INFRA-RED

BALANCE GAS NITROGEN
ANALYZER READINGS: Z = Zero Gas T = Test Gas R = Reference Gas

Component CARBON MONOXIDE

First Analysis Date 05/05/92 Units mv

Z	00.00	R	47.8	T	31.9
R	47.8	Z	00.02	T	31.9
Z	-00.03	T	31.9	R	47.8

Mean Test Assay 16.17 PPM

Second Analysis Date 05/14/92 Units mv

Z	00.00	R	47.8	T	31.9
R	47.8	Z	00.04	T	31.9
Z	-00.02	T	31.9	R	47.8

Mean Test Assay 16.17 PPM

Component _____

Date _____ Units mv

Z	_____	R	_____	T	_____
R	_____	Z	_____	T	_____
Z	_____	T	_____	R	_____

Mean Test Assay _____

Date _____ Units mv

Z	_____	R	_____	T	_____
R	_____	Z	_____	T	_____
Z	_____	T	_____	R	_____

Mean Test Assay _____

Component _____

Date _____ Units mv

Z	_____	R	_____	T	_____
R	_____	Z	_____	T	_____
Z	_____	T	_____	R	_____

Mean Test Assay _____

Date _____ Units mv

Z	_____	R	_____	T	_____
R	_____	Z	_____	T	_____
Z	_____	T	_____	R	_____

Mean Test Assay _____

Chronology: Date _____
Assay _____

Analyst M. JOHNSON

Approved By: A.F. Lange
A.F. LANGE, Ph.D., MANAGER, QA



Scott Specialty Gases, Inc.

Shipped from: 2600 CAJON BLVD., SAN BERNARDINO CA 92411
 Phone: 714-887-2571 Fax: 714-887-0549

C E R T I F I C A T E O F A N A L Y S I S

CARNOT PROJECT #: 02-22304
 ATTN: RICK MADRIGAL POW: 8957
 15991 RED HILL AVE ITEM #: 020203013601A1
 SUITE 110 DATE: 10/29/92
 MUSTIN CA 92680

CYLINDER #: AA15966 ANALYTICAL ACCURACY: +/-1%

FILL PRESSURE: 1950 PSIG

COMPONENT	REQUESTED GAS		ANALYSIS	
	COND. MOLES		(MOLES)	
CARBON DIOXIDE	20.5	PCT	22.74	PCT
OXYGEN	0.	PCT	8.003	PCT
NITROGEN		BAL.		BAL.

CARBON DIOXIDE
 CERTIFIED TO HAVE BEEN ANALYZED
 AND VERIFIED TO BE CORRECT BY
 ANALYSIS.

ANALYZED MASTER GAS
 ANALYST MUST CERTIFY ALL THIS
 INDEPENDENT LABORATORY

ANALYST: JDLT
 JOSEPH DE LA TORRE

APPROVED BY: [Signature]
 DE LA TORRE



Scott Specialty Gases, Inc.

Shipped From: 2400 CAJON BLVD.,
 SAN BERNARDINO CA 92411
 Phone: 714-887-2571 Fax: 714-887-0549

C E R T I F I C A T E O F A N A L Y S I S

CARNOT
 ATTN: RICK MADRIGAL
 15991 RED HILL AVE
 SUITE 110
 TUSTIN CA 92680

PROJECT #: 02-22304
 PO#: 8957
 ITEM #: 020203000701AL
 DATE: 10/26/92

CYLINDER #: AL9006709 ANALYTICAL ACCURACY: 47-12
 FILL PRESSURE: 2000 PSIG

COMPONENT	REQUESTED GAS		ANALYSIS	
	CONC MOLES		(MOLES)	
CARBON DIOXIDE	15.0	PCT	15.15	PCT
OXYGEN	5.0	PCT	5.160	PCT
NITROGEN		BAL.		BAL.

CARNOT 11/25 BINES
 CERTIFIED TO HAVE BEEN BLENDED
 AND VERIFIED TO BE CORRECT BY
 ANALYSIS.

GRANDFIELD CARTER GAS -
 AGAINST MOST CERTIFIED BATCHES
 INDEPENDENT LABORATORY

ANALYST: MTJ APPROVED BY: [Signature]
 RAYMOND JOHNSON DEBORAH JOHNSON

Appendix B.4
Chain of Custody

CERTIFICATION OF SAMPLE RECEIPT

PROJECT # 54700 OUTSIDE LAB# Y SAMPLE DATE 12, 30, 92
 SAMPLE LOCATION In & Out METHOD(S) 25.2 PROJECT MANAGER Craig F...
 COMPLIANCE TEST (Y/N) Y DATE DUE _____

TEST #	COMPLETE DESCRIPTION
1-Hc-In	Tedlar Bag Samples
1-Hc-Out	SCAOMD Method 25.2 Analysis
2-Hc-In	
2-Hc-Out	In - Landfill Gas
	Out - Boiler Exhaust

CHAIN OF CUSTODY PRIOR TO SHIPMENT:

RELEASED BY	DATE	TIME	RECEIVED BY	DATE	TIME
<i>[Signature]</i>	12/30/92	3:21		1 1	:
	1 1	:		1 1	:
	1 1	:		1 1	:

OUTSIDE LAB INFORMATION

SAMPLES SHIPPED TO Atm AA Inc. CARNOT SHIPPER Reslit at Way
 _____ DATE 1 1
 _____ CARRIER # _____
 PHONE: () - _____ AIR BILL # _____
 RECIPIENT _____
 COMPANY _____
 DATE 1 1



CHAIN OF CUSTODY

PROJECT # 54700 OUTSIDE LAB REQUIRED (Y/N) N SAMPLE DATE 12 1301 92
 CLIENT/LOCATION PVERG PROJECT MANAGER Craig Fry
 SAMPLE LOCATION Stack METHOD(S) S.I TECHNICIAN _____
 COMPLIANCE TEST (Y/N) Y DATE DUE 1/8/92

DATE	TIME	TEST #	SAMPLE DESCRIPTION	CONTAINERS	SAMPLER	COMMENTS
12/30	:	1-PM-PV	Method S.I	2	CF	
12/30	:	2-PM-PV	" " S.I	2	CF	

TRANSFER OF SAMPLES FROM FIELD SAMPLE CUSTODIAN (FSC):

RELEASED BY (FSC)	DATE AND TIME	RECEIVED BY	DATE AND TIME
<i>[Signature]</i>	12/30/92 3:45	<i>Marc Kodabangh</i>	12/30/92 15:45

ANALYSIS REQUIRED: SCAQD Method S.I w/ Organic Extracts



CHAIN OF CUSTODY FORM

FIELD PERSONNEL

SAMPLE NO.	1	2	3	4	5	6	7
SAMPLE TYPE *	STGS						
SAMPLING DATE	12/30/92						
LANDFILL	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF
SUBLOCATION	-						
OVA READING -- OR -- CH4/O2	0.5						
BAG NO.	796						
COMMENTS	PVLF Boiler Emissions Testing						

SAMPLE TAKEN BY: Theresa [Signature]
 DELIVERED BY: [Signature]

DATE: Dec/30/92
 DATE: Dec/30/92

LAB PERSONNEL

SAMPLE NO.	1	2	3	4	5	6	7
LOG NO.	95194						
BAG COND. **	F						
COMMENTS							

SAMPLE RECEIVED BY: Willam [Signature] DATE: 12/30/92

- *AIR - AMBIENT AIR
- ILSS - SURFACE GAS
- LFGS - LANDFILL GAS (RAW)
- STGS - FLARE EXHAUST
- LPGP - BOUNDARY PROBE
- ** F = MORE THAN 1/2 FULL
- L = 1/4 TO 1/2 FULL
- E = LESS THAN 1/4 FULL BUT CONTAINS SOME SAMPLE
- N = NO SAMPLE IN BAG

RETURN THIS FORM TO MOON CHUNG AFTER SAMPLES ARE RECEIVED BY LAB PERSONNEL.

CHAIN OF CUSTODY FORM

FIELD PERSONNEL

SAMPLE NO.	1	2	3*	4	5	6	7
SAMPLE TYPE *	LFGS	LFGS	STGS				
SAMPLING DATE	12/30	12/30	12/30				
LANDFILL	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF	2-PVLF
SUBLOCATION	—	—	—				
OVA READING --- OR --- CH4/O2	— 20/11	— 20/11	— 0/5	---	---	---	---
BAG NO.	612	512	797				
COMMENTS	OVER Emissions Testing						

SAMPLE TAKEN BY : *Thomas A. Ellis* * Sample #3 obtained from Carnot DATE : 12/30/92
2-11-93 EDWARDS
 DELIVERED BY : *Thomas A. Ellis #92* DATE : 12/30/92

LAB PERSONNEL

SAMPLE NO.	1	2	3	4	5	6	7
LOG NO.	95191	95192	95192				
BAG COND. **	F	F	F				
COMMENTS							

SAMPLE RECEIVED BY : *Will A. Kelly* DATE : 12/30/92

- *AIR - AMBIENT AIR
 - ILSS - SURFACE GAS
 - LFGS - LANDFILL GAS (RAW)
 - STGS - FLARE EXHAUST
 - LPGP - BOUNDARY PROBE
- ** F = MORE THAN 1/2 FULL
 - L = 1/4 TO 1/2 FULL
 - E = LESS THAN 1/4 FULL BUT CONTAINS SOME SAMPLE
 - N = NO SAMPLE IN BAG

RETURN THIS FORM TO *Moon Chung* AFTER SAMPLES ARE RECEIVED BY LAB PERSONNEL.

APPENDIX C
FIELD AND LAB DATA SUMMARIES


Appendix C.1
Sample Locations

CARNOT
SAMPLING POINT LOCATION DATA - EPA METHOD 1

PLANT: PV Unit 1 DATA BY: RM

DATE: 12/30/92

TEST LOCATION: Stack



UPSTREAM DIST./DIA.: 1.58
 DOWNSTREAM DIST./DIA.: 6.44
 COUPLING LENGTH: 5.5"
 NO. OF SAMPLING PTS.: 16
 STACK DIMENSION: 41"
 STACK AREA, FT²: 9.168

SAMPLE POINT	% OF DIAMETER	IN. FROM NEAR WALL	IN. FROM NOZZLE*
1	3.2	1.3	6.8
2	10.5	4.3	9.8
3	19.4	7.95	13.5
4	32.3	13.2	18.7
5	67.7	27.8	33.3
6	80.6	33.0	38.5
7	89.5	36.7	42.2
8	96.8	39.7	45.2

*INCHES FROM WALL PLUS COUPLING LENGTH

CEM System Bias and Linearity Correction Calculations						
Test No.	O2	CO2	CO	NOx	Acceptance Criteria	Run Status
1-CEM-PV						
Linearity						
Analyzer Range	10	25	100	25		
High Cal	8.01	22.74	16.17	22.81		
Low Cal	5.16	15.15	11.22	12.68		
Analyzer Reads	5.22	15.35	11.40	12.50		
Analyzer Cal. Error, %	0.6	0.8	0.2	-0.7	< 2%	PASS
System Bias						
Pretest Bias						
Zero	0.02	0.00	0.00	0.00	< 5%	PASS
Span	5.14	15.20	11.16	12.55	< 5%	PASS
Posttest Bias						
Zero	0.02	0.08	0.00	0.00	< 5%	PASS
Span	5.30	14.95	10.70	12.50	< 5%	PASS
Span Value	5.16	15.15	11.22	12.68		
Zero Drift, %	0.00	0.32	0.00	0.00	< 3%	PASS
Span Drift, %	1.60	-1.00	-0.46	-0.20	< 3%	PASS
Test Ave.	5.33	12.95	11.67	18.18		
Corrected Ave.	5.27	13.01	11.98	18.40		

CEM System Bias and Linearity Correction Calculations

Test No.	O2	CO2	CO	NOx	Acceptance Criteria	Run Status
2-CEM-PV						
Linearity						
Analyzer Range	10	25	100	25		
High Cal	8.01	22.74	16.17	22.81		
Low Cal	5.16	15.15	11.22	12.68		
Analyzer Reads	5.22	15.35	11.40	12.50		
Analyzer Cal. Error, %	0.6	0.8	0.2	-0.7	< 2%	PASS
System Bias						
Pretest Bias						
Zero	0.02	0.05	0.00	0.05	< 5%	PASS
Span	5.14	15.10	10.80	12.60	< 5%	PASS
Posttest Bias						
Zero	0.10	0.10	0.00	0.05	< 5%	PASS
Span	5.13	15.30	10.00	12.40	< 5%	PASS
Span Value	5.16	15.15	11.22	12.68		
Zero Drift, %	0.80	0.20	0.00	0.00	< 3%	PASS
Span Drift, %	-0.10	0.80	-0.80	-0.80	< 3%	PASS
Test Ave.	5.36	12.88	14.67	18.02		
Corrected Ave.	5.39	12.83	15.83	18.30		

CARNOT CEM PERFORMANCE DATA

CLIENT/LOCATION: PVERG DATE: 12-30-92
 BY: CHF

SYSTEM CONFIGURATION				
ANALYZERS IN SERVICE				
ANALYZERS:	O ₂	CO ₂	CO	NOx
MODEL:	Teledyne	Horiba	Teco	Teco 10
SERIAL NO.:			48	
PROBE		SAMPLE CONDITIONER		
LENGTH:	4'	CONDENSER-VACUUM SIDE (CHECK FLOW): <input checked="" type="checkbox"/>		
LINER MATERIAL:	SS	CONDENSER-PRESSURE SIDE (CHECK FLOW): <input checked="" type="checkbox"/>		
HEATED PROBE (Y/N):	N	CONDENSER TEMPERATURE: <u>34</u>		
HEATED LINE (Y/N):	Y	FILTER CONDITION (COND. OR DATE LAST CHANGED): <u>12-10-92</u>		
SAMPLE LINE		SYSTEM LEAK CHECK		
LENGTH:	100	PRE-TEST (cfh):		
LINER MATERIAL:	Teflon	POST-TEST (cfh):		
SYSTEM BIAS LINE:	Poly	LEAK RATE (%) = $\frac{\text{POST-TEST (cfh)}}{\text{SYSTEM FLOW RATE (cfm)} \times 60} \times 100 = \underline{\hspace{2cm}}\%$		
ON-STACK CONDITIONER		NOx CONVERSION EFFICIENCY		
IN SERVICE (Y/N):	Y	HIGH CAL NOx <u>22.81</u>		
KNOCK-OUT CONDITION (CHECK FLOW):		HIGH CAL NO (AS FOUND) <u>22.357 22.43</u>		
COOLANT:	Ice	LOW CAL NOx _____		
		LOW CAL NO (AS FOUND) <u>1</u>		
OPERATING CONDITIONS				
SAMPLE PRESSURE:	5	SYSTEM RESPONSE TIME CHECK		
SAMPLE VACUUM:		UPSCALE: _____ sec.		
NOx VACUUM:	29	DOWNSCALE: _____ sec.		

REFERENCE METHOD GASEOUS MEASUREMENTS

CLIENT/LOCATION: PVERG

CONDITION: _____

DATE: 12-30-92

OPERATOR: CHF

TEST NUMBER: 2-CEM-PV

TEST LOCATION: _____

TEST NO.	TIME		SAMPLE POINT/ CONDITION	DRY, UNCORRECTED						CORRECTED TO % DRY		
				O ₂	CO ₂	CO	NOx	NO	NO ₂	SO ₂	CO	NOx
SPAN GAS CONCENTRATION												
			AS FOUND ANAL. SPAN									
			SYSTEM ZERO	0.02	0.05	0.0	0.05					
			SYSTEM SPAN	5.14	15.1	10.8	12.6					
	BEGIN	1122	Start	5.4	12.9	12	18.0					
	END	1132										
	BEGIN	1132		5.35	12.9	12	18.2					
	END	1142										
	BEGIN	1142		5.35	12.9	15	18.1					
	END	1152										
	BEGIN	---	Port Change									
	BEGIN	1200	New Port	5.3	12.8	14	18.3					
	END	1210										
	BEGIN	1210		5.35	12.9	15	17.9					
	END	1220										
	BEGIN	1220	End	5.4	12.9	20	17.6					
	END	1230										
	BEGIN	---										
	END	---										
	BEGIN	---										
	END	---										
	BEGIN	---										
	END	---										
	BEGIN	---										
	END	---										
			SYSTEM ZERO	0.1	0.1	0.0	0.05					
			SYSTEM SPAN	5.13	15.13	10.0	12.4					
RAW AVERAGE				5.36	12.88	14.67	18.02					
CORRECTED AVERAGE				✓	✓	✓	✓					
COMMENTS:												

REFERENCE METHOD GASEOUS MEASUREMENTS

CLIENT/LOCATION: PVERG

CONDITION: _____

DATE: 12-30-92

OPERATOR: CHF

TEST NUMBER: 1-CFM-PV


TEST LOCATION: Stack

TEST NO.	TIME		SAMPLE POINT/ CONDITION	DRY, UNCORRECTED						CORRECTED TO % DRY			
				O ₂	CO ₂	CO	NO _x	NO	NO ₂	SO ₂	CO	NO _x	SO ₂
SPAN GAS CONCENTRATION					60.60	56.10	50.72						
			AS FOUND ANAL SPAN	5.162	15.15	11.22	17.68						
			SYSTEM ZERO	.02	0.0	0.0	0.0						
			SYSTEM SPAN	5.14	15.2	11.16	12.55						
	BEGIN	435	Start	5.35	13.0	10.0	17.8						
	END	945											
	BEGIN	945		5.35	12.9	10	17.8						
	END	955											
	BEGIN	955	Port Change	5.4	12.9	12	17.8						
	END	1005											
	BEGIN	1015	Next Port	5.3	13.0	12	18.3						
	END	1025											
	BEGIN	25		5.35	12.9	14	18.6						
	END	35											
	BEGIN	35		5.25	13.0	12	18.8						
	END	1045											
	BEGIN	---											
	END	---											
	BEGIN	---											
	END	---											
	BEGIN	---											
	END	---											
	BEGIN	---											
	END	---											
	BEGIN	---											
	END	---											
	BEGIN	---											
	END	---											
			SYSTEM ZERO	.02	.08	0.0	0.0						
			SYSTEM SPAN	5.3	14.95	10.7	12.5						
RAW AVERAGE				5.33	12.95	11.67	18.18						
CORRECTED AVERAGE				✓	✓	✓	✓						
COMMENTS:													

CARNOT
CO Range switched from 0-20 to 0-100 before start of Test

MOC-3e-10-02 CARTRIDGE DETECTOR

08001000331712 21



12.55 / 12.68 NO_x

11.16 / 11.22 CO

15.2 / 15.15 CO₂

9.14 / 9.16 O₂

Analytical Span 8.59

NO_x 12.5 / 12.68

CO 11.34 / 11.22

CO 16.17

NO_x 27.81

15.35 / 15.15 CO₂

9.22 / 9.16 O₂

Linearity Set 1. 8.01=0


27.74 CO

PVER G 12-30-92

IF J11EED007910

5808-5023717 31

5808-5023717 31



10:25

10:15
Next 10:05

Port Change

10:05

9:55

CHART NO. 11-25-201

9:45

Start 935
Test. 1-CEM-PV

CO = 0-100

12:55 / 12:68 NOx

11:16 / 11:22 CO

1142

1132

1122

Stat Test 2-CEM-PA

CHART NO. 11

12.5 / 12.69 NOx

10.8 / 11.22 CO

15.1 / 15.15 CO2

5.14 / 5.16 O2

System Span

11.18 / 11.22 CO

16.17 SO2

CO Linearity 0-100 Range

12.5 / 12.69 NOx

10.68 / 11.22 CO

14.95 / 15.15 CO2

5.3 / 5.16 O2

1049 Lm

1045

035


DATE: 11/22/11

18 011 11 22 011 11 22

035

San Fernando, CA 91340

CHART NO. 4110115-20M



12.4 / 12.63 NOX

5.13 / 5.16 O₂

15.17 / 15.18 CO

10.0 / 11.22 CO

System Proc

1230 End

1220

1210

1200 Start

Port. Change

1152

1142

100
200
300
400
500
600
700
800
900
1000

Appendix C.3
Particulate and Calculated Exhaust Flow Rate Data

CARNOT

SOURCE TEST DATA SUMMARY

Client/Location.....	PVERG	*	Date.....	12/30/92
Test Number.....	1-PM-PV	*	Data By.....	CHF
Test Method.....	5.1	*	Sample Location.....	EXHAUST
Fuel.....	LFG	*	Reference Temp (F).....	60
Sample Train.....	ES-42	*	Unit.....	1
Pitot Factor	0.840	*	Meter Cal Factor.....	0.9990
Stack Area (sq ft).....	9.168	*	Sample Time (Min).....	64
Bar Press (in Hg).....	29.96	*	Nozzle Diam (in).....	0.247
Meter Vol (acf).....	32.939	*	Meter Temp (F).....	71.4
Stack Press (iwg).....	0.145	*	Stack Temp (F).....	335.7
Vel Head (iwg).....	0.46790	*	Stack O2 (%).....	5.27
Liquid Vol (ml).....	117.4	*	Stack CO2 (%).....	13.01
Meter Press (iwg).....	0.923	*	Start/Stop Time.....	0935/1047
Std Sample Vol (SCF).....				32.314
Moisture Fraction.....				0.144
Stack Gas Mol Wt.....				28.52
Stack Gas Velocity (ft/sec).....				47.32
Stack Flow Rate (wacfm).....				26032
Stack Flow Rate (dscfm).....				14576
Isokinetic Ratio (%).....				95.41
Particulate Catch, mg.....				7.10
Particulate Catch, (organic fraction) mg.....				0.00
Grain Loading, gr/dscf.....				0.0034
Grain Loading, (organic fraction) gr/dscf.....				0.0000
Grain Loading, gr/dscf @ 12% CO2.....				0.0031
Part Emission lb/hr.....				0.4236

CARNOT

SOURCE TEST DATA SUMMARY

Client/Location.....	PVERG	*	Date.....	12/30/92
Test Number.....	2-PM-PV	*	Data By.....	CHF
Test Method.....	5.1	*	Sample Location.....	EXHAUST
Fuel.....	LFG	*	Reference Temp (F).....	60
Sample Train.....	ES-42	*	Unit.....	1
Pitot Factor	0.840	*	Meter Cal Factor.....	0.9990
Stack Area (sq ft).....	9.168	*	Sample Time (Min).....	64
Bar Press (in Hg).....	29.96	*	Nozzle Diam (in).....	0.247
Meter Vol (acf).....	33.677	*	Meter Temp (F).....	79.8
Stack Press (iwg).....	0.118	*	Stack Temp (F).....	336.8
Vel Head (iwg).....	0.47500	*	Stack O2 (%).....	5.39
Liquid Vol (ml).....	120.0	*	Stack CO2 (%).....	12.83
Meter Press (iwg).....	0.955	*	Start/Stop Time.....	1122/1232
Std Sample Vol (SCF).....				32.526
Moisture Fraction.....				0.146
Stack Gas Mol Wt.....				28.47
Stack Gas Velocity (ft/sec).....				47.75
Stack Flow Rate (wacfm).....				26268
Stack Flow Rate (dscfm).....				14654
Isokinetic Ratio (%).....				95.52
Particulate Catch, mg.....				4.30
Particulate Catch, (organic fraction) mg.....				0.00
Grain Loading, gr/dscf.....				0.0020
Grain Loading, (organic fraction) gr/dscf.....				0.0000
Grain Loading, gr/dscf @ 12% CO2.....				0.0019
Part Emission lb/hr.....				0.2562

CARNOT

LACSD

SAMPLE TRAIN TEST DATA

CLIENT PV Boiler Unit / UNIT 1 TEST NO. 1-PM-PV METHOD S.1 PAGE 1 OF 1
SAMPLE LOCATION Stack #1 TEST CONDITION AMB. TEMP., °F PROJECT # 54720
OPERATOR/ASSISTANT DE/RM METER VOL. (START/END) 1 DATE 12/30/92

PRE-TEST DATA:

Barometric Press., in. Hg. 29.96
Assumed Stack Press.
Assumed Moisture 15.9
Assumed Molecular Wt.
Assumed ΔH
Stack Diameter, in.
Sample Time: Total 64
per point 4
Total of Traverso Points 16

EQUIPMENT INFO:

Motor No. 42
Motor, Yd. 499
CFM @ ΔH = 1.0 553
Pilot ID, Cp 34
O2/CO, Method Y
Teflon Connecting Y
Lino (Y/N)
Probe: Mat'l 4'
Length 21955
Nozzle: Mat'l
Diam. 247
Filter: No. 63-706
Mat'l Glass

Table with 5 columns: Imp., Mat'l, Wt.(End), Wt.(Start), Wt.(g). Rows #1-#5 showing H2O and other measurements.

SAMPLE TRAIN LEAK CHECK:

Table with 4 columns: CFM, Vac., Pilot, Init. Rows for Pre-Test and Post-Test.

PRE-TEST CALIBRATION CHECK:

Table with 4 columns: Time, ΔH, Meter Reading, Meter Temp. In, Out.

ΔH = 2.06 x ΔP
1.96

Main data table with columns: SAMPLE POINT, TIME, METER CONDITIONS (ΔP, ΔH, METER READING, STACK, PROBE), TEMPERATURES (METER IN, OUT, OVEN, IMP. OUT), O2, VAC., STATIC PRESS. (iwg), CHAIN OF CUSTODY INFORMATION.

CARNOT

SAMPLE TRAIN TEST DATA

CLIENT LACSD - PV UNIT # 1 TEST NO. 2-PM-PV METHOD S.1 PAGE 1 OF 1
 SAMPLE LOCATION Baker #1 Stack TEST CONDITION _____ AMB. TEMP., °F _____ PROJECT # _____
 OPERATOR/ASSISTANT DE/RH METER VOL. (START/END) 1 DATE 12/30/92

PRE-TEST DATA:

Barometric Press., In. Hg. 29.96
 Assumed Stack Press. .1
 Assumed Moisture 15.9
 Assumed Molecular Wt. _____
 Assumed ΔP _____
 Assumed ΔH _____
 Stack Diameter, in. 41"
 Sample Time: Total 4
 per point 4
 Total of Traverso Points 16

EQUIPMENT INFO:

Motor No. 42
 Motor, Yd. 1.999
 CFM @ ΔH = 1.0 .553
 Pilot ID, Cp .84
 O₂/CO₂ Method Y
 Teflon Connecting _____
 Line (Y/N) _____
 Probe: Mat'l Glass
 Length 42
 Nozzle: Mat'l 1.247
 Diam. Glass
 Filter: No. 63-707
 Mat'l _____

Imp.	Mat'l	Wt.(End)	Wt.(Start)	Wt.(g)
#1	H ₂ O	685.8	575.4	110.4
#2	H ₂ O	604.8	603.2	1.6
#3	Ø	491.9	491.7	0.2
#4	SO ₂	769.2	761.4	7.8
#5				

SAMPLE TRAIN LEAK CHECK:

	CFM	Vac.	Pilot	Init.
Pro-Test	<u>.008</u>	<u>15</u>	<u>✓</u>	<u>DE</u>
Post-Test	<u>.008</u>	<u>15</u>	<u>✓</u>	<u>DE</u>

PRE-TEST CALIBRATION CHECK:

	Time	ΔH	Meter Reading	Meter Temp. In	Meter Temp. Out
Init.	_____	_____	_____	_____	_____
Final	_____	_____	_____	_____	_____

Total POST TEST INFO:
 Filter Appearance _____
 Impinger Appearance _____
 Silica Gel Spent (Y/N) _____

ΔH = 1.96 x ΔP
1.98 2.05

SAMPLE POINT	TIME	METER CONDITIONS		METER READING	STACK	PROBE	TEMPERATURES, °F			IMP. OUT	O ₂	VAC.	STATIC PRESS. Iwg	CHAIN OF CUSTODY INFORMATION
		ΔP	ΔH				METER IN	METER OUT	OVEN					
NE 8	1102	.55	1.08	608.605	338		74	73		160		4	Impingers Loaded	CHF
7	06	.52	1.01	610.900	339		79	74		"		3.8	Impingers Recovered	
6	30	.54	1.06	613.011	338		79	74		"		4 +.15	Filter Loaded	CHF
5	34	.59	1.16	615.400	338		82	75		53		4.1	Filter Recovered	
4	38	.55	1.08	617.531	337		83	76		"		4.0	Probe Wash	
3	42	.48	.94	619.785	335		84	77		"		3.5		TEST SUMMARY
2	46	.39	.76	621.935	335		84	77		"		3	Calculated by:	DE
1	50	.36	.71	623.777	334		85	78		"		3	Checked by:	DE CHF ✓
-	54			625.564									Sample Vol., c.i.	33.677 ✓
SE 8	1200	.50	.99	625.564	339		81	78		"		3.5	Stack Press., iwg	+1.181 ✓
7	04	.51	1.01	627.800	339		83	78		"		3.8	ΔH, iwg	9467.955 ✓
6	08	.54	1.11	629.911	338		85	78		"		4	ΔP, iwg	.4750 ✓
5	12	.59	1.20	632.211	338		85	78		"		4.1	Motor Temp., °F	79.78 ✓
4	16	.48	.98	634.535	334		85	78		"		3.8	Stack Temp., °F	336.75 ✓
3	20	.41	.84	636.695	335		85	78		"		3.5	Water Collected, g	120.0 ✓
2	24	.38	.78	638.700	336		85	78		"		3.2 +.09	O ₂ /CO ₂	
1	28	.28	.57	640.511	335		85	79		"			Comments:	
-	32			642.282										

PARTICULATE
SCAQMD 5.1, WET IMPINGEMENT

PROJECT # 54700
 CLIENT/LOCATION PVERG
 SAMPLE LOCATION STACIC
 TEST # 1-PM-PV

#5383
 WATER BLANK mg/ml 0.0017 mg/ml
 #5384
 MeCl₂ BLANK mg/ml 0.0056 mg/ml

SAMPLE DATE 12/30/92
 DATA ANALYSIS DATE 1/4/93
 ANALYST INITIALS ML

ITEM	NUMBER	FINAL WEIGHT (g)	TARE WEIGHT (g)	GAIN WEIGHT (mg)	BLANK CORRECTION (mg)	ALLOQUOT CORRECTION (ml/ml)	NET GAIN (mg/SAMPLE)
1. GLASS FIBER FILTER (BACKUP FILTER)	63-706	0.1995	0.1992	0.3	-	-/-	0.3
2. INSOLUBLE SOLIDS FILTER	63-708	0.1971	0.1974	0.0	-	TV/TV 700/700	0.0
3. IMPINGERS a. WATER FRACTION	5385	28.5150	28.5110	4.0	0.6	700/350	6.8
b. ORGANIC EXTRACTION <u>125 ml MeCl₂</u>	5386	28.3449	28.3442	0.7	0.7	700/350	0.0
						/	.
H ₂ O BLANK	5383	31.8716	31.8710	0.6	-	-/350	.
MeCl ₂ EXTRACTION BLANK	5384	28.5008	28.5001	0.7	-	-/125	.

TOTAL PARTICULATE = 7.1 mg

METHOD OF SAMPLE PREP/NOTES

**PARTICULATE
SCAQMD 5.1, WET IMPINGEMENT**

PROJECT # 54700
 CLIENT/LOCATION PYERG
 SAMPLE LOCATION STACK
 TEST # 2-PM-PV

#5383
 WATER BLANK mg/ml 0.0017 mg/ml
 #5384
 MeCl₂ BLANK mg/ml 0.0056 mg/ml

SAMPLE DATE 12/30/92
 DATA ANALYSIS DATE 1/4/93
 ANALYST INITIALS ML

ITEM	NUMBER	FINAL WEIGHT (g)	TARE WEIGHT (g)	GAIN WEIGHT (mg)	BLANK CORRECTION (mg)	ALIQOT CORRECTION (ml/ml)	NET GAIN (mg SAMPLE)
1. GLASS FIBER FILTER (BACKUP FILTER)	63-707	0.1983	0.1978	0.5	-	-/-	0.5
2. INSOLUBLE SOLIDS FILTER	63-709	0.1971	0.1973	0.0	-	TVTV 720/ 720	0.0
3. IMPINGERS a. WATER FRACTION	5387	31.1122	31.097	2.5	0.6	720/ 360	3.8
b. ORGANIC EXTRACTION <u>125 ml MeCl₂</u>	5388	28.6491	28.6484	0.7	0.7	720/ 360	0.0
						/	.
						/	.
						/	.

TOTAL PARTICULATE = 4.3 mg

METHOD OF SAMPLE PREP/NOTES

Appendix C.4
Inlet Moisture and Calculated Flow Rate Data

SOURCE TEST DATA SUMMARY

Client/Location.....	PVERG	*	Date.....	12/30/92
Test Number.....	1-H2O-PV	*	Data By.....	CHF
Test Method.....	4.1	*	Sample Location.....	INLET
Fuel.....	LFG	*	Reference Temp (F).....	60
Sample Train.....	ES-28	*	Unit.....	1
Pitot Factor	N/A	*	Meter Cal Factor.....	1.0000
Stack Area (sq ft).....	N/A	*	Sample Time (Min).....	30
Bar Press (in Hg).....	29.96	*	Nozzle Diam (in).....	N/A
<hr/>				
Meter Vol (acf).....	23.808	*	Meter Temp (F).....	77.67
Stack Press (iwg).....	N/A	*	Stack Temp (F).....	N/A
Vel Head (iwg).....	N/A	*	Duct O2 (%).....	11.20
Liquid Vol (ml).....	19.60	*	Duct CO2 (%).....	15.80
Meter Press (iwg).....	2.00	*	Duct CH4 (%).....	19.20
		*	Start/Stop Time.....	0952/1022
<hr/>				
Std Sample Vol (SCF).....				23.15
Moisture Fraction.....				0.038
Stack Gas Mol Wt.....				28.27
Stack Gas Velocity (ft/sec).....				0.00
Stack Flow Rate (scfm).....				3730
Stack Flow Rate (dscfm).....				3589

SOURCE TEST DATA SUMMARY

Client/Location.....	PVERG	*	Date.....	12/30/92
Test Number.....	2-H2O-PV	*	Data By.....	CHF
Test Method.....	4.1	*	Sample Location.....	INLET
Fuel.....	LFG	*	Reference Temp (F).....	60
Sample Train.....	ES-28	*	Unit.....	1
Pitot Factor	N/A	*	Meter Cal Factor.....	1.0000
Stack Area (sq ft).....	N/A	*	Sample Time (Min).....	35
Bar Press (in Hg).....	29.96	*	Nozzle Diam (in).....	N/A
<hr/>				
Meter Vol (acf).....	25.170	*	Meter Temp (F).....	68.00
Stack Press (iwg).....	N/A	*	Stack Temp (F).....	N/A
Vel Head (iwg).....	N/A	*	Duct O2 (%).....	11.20
Liquid Vol (ml).....	23.60	*	Duct CO2 (%).....	16.10
Meter Press (iwg).....	2.00	*	Duct CH4 (%).....	19.90
		*	Start/Stop Time.....	1130/1205
<hr/>				
Std Sample Vol (SCF).....				24.93
Moisture Fraction.....				0.042
Stack Gas Mol Wt.....				28.19
Stack Gas Velocity (ft/sec).....				0.00
Stack Flow Rate (scfm).....				3714
Stack Flow Rate (dscfm).....				3557

CARNOT

SAMPLE TRAIN TEST DATA

CLIENT PVERG UNIT 1 TEST NO. 1-110-PV METHOD 4.1 PAGE 1 OF 1
 SAMPLE LOCATION Inlet TEST CONDITION _____ AMB. TEMP., °F _____ PROJECT # _____
 OPERATOR/ASSISTANT CHF METER VOL. (START/END) 1 DATE 12/30/77

PRE-TEST DATA:
 Barometric Press., In. Hg. 29.96
 Assumed Stack Press. _____
 Assumed Moisture _____
 Assumed Molecular Wt. _____
 Assumed ΔP _____
 Assumed ΔH _____
 Stack Diameter, In. _____
 Sample Time: Total _____
 per point _____
 Total of Traverse Points _____
 ΔH = _____ x ΔP

EQUIPMENT INFO:
 Motor No. ES-78
 Motor, Yd. LOO
 CFM @ ΔH = 1.0 _____
 Pilot ID, Cp _____
 O₂/CO₂ Method _____
 Tufflon Connecting _____
 Line (Y/N) _____
 Probe: Mat'l _____
 Length _____
 Nozzle: Mat'l _____
 Diam. _____
 Filter: No. _____
 Mat'l _____

Imp.	Mat'l	Wt.(End)	Wt.(Start)	Wt.(g)
#1	H ₂ O	625.8	614.6	11.2
#2	H ₂ O	614.8	613.5	1.3
#3	Ø	515.3	514.6	0.7
#4	SG	779.9	773.5	6.4
#5				

Total _____
POST TEST INFO:
 Filter Appearance _____
 Impinger Appearance _____
 Silica Gel Spent (Y/N) _____

SAMPLE TRAIN LEAK CHECK:

	CEM	Vac.	Pilot	Init.
Pre-Test	.001	10		CHF
Post-Test	.001	10		CHF

PRE-TEST CALIBRATION CHECK:

	Time	ΔH	Motor Reading	Motor Temp. In	Motor Temp. Out
Init.					
Final					

SAMPLE POINT	TIME	METER CONDITIONS			TEMPERATURES, °F							STATIC PRESS. Iwg	CHAIN OF CUSTODY INFORMATION	
		ΔP	ΔH	METER READING	STACK	PROBE	METER		OVEN	IMP. OUT	O ₂			VAC.
							IN	OUT						
	952		2.0	145.307			79	72		48	-	0	Impingers Loaded	
	1002		2.0	153.75			84	74		52		0	Impingers Recovered	
	1012		2.0	160.25			83	74		52		0	Filter Loaded	
	1022		2.0	169.115									Filter Recovered	
													Probe Wash	
													TEST SUMMARY	
													Calculated by: CHF	
													Checked by: CHF ✓	
													Sample Vol., c.f. 23.808 ✓	
													Stack Press., iwg	
													ΔH, Iwg 2.0 ✓	
													ΔP, iwg	
													Motor Temp., °F 77.67 ✓	
													Stack Temp., °F	
													Water Collected, g 19.6 ✓	
													O ₂ /CO ₂	
													Comments:	

CARNOT

PVER6

SAMPLE TRAIN TEST DATA

CLIENT PVER6 UNIT 1 TEST NO. 2-Ho-In METHOD 4.1 PAGE 1 OF 1
 SAMPLE LOCATION Inlet TEST CONDITION _____ AMB. TEMP., °F _____ PROJECT # _____
 OPERATOR/ASSISTANT CHF METER VOL. (START/END) 1 DATE 12-30-92

PRE-TEST DATA: Barometric Press., In. Hg. <u>29.46</u> Assumed Stack Press. _____ Assumed Moisture _____ Assumed Molecular Wt. _____ Assumed ΔP _____ Assumed ΔH _____ Stack Diameter, In. _____ Sample Time: Total _____ per point _____ Total of Traversal Points _____ ΔH = _____ x ΔP	EQUIPMENT INFO: Meter No. <u>ES-28</u> Meter, Yd. <u>1.000</u> CFM @ ΔH = 1.0 _____ Pitot ID, Cp _____ O ₂ /CO ₂ Method _____ Tollen Connecting _____ Uno (Y/N) _____ Probo: Mat'l _____ Length _____ Nozzle: Mat'l _____ Diam. _____ Filter: No. _____ Mat'l _____	Imp.	Mat'l	Wt.(End)	Wt.(Start)	Wt.(g)	SAMPLE TRAIN LEAK CHECK:						
		#1	H ₂ O	640.8	625.8	15.0	CFM	Vac.	Pitot	Init.			
		#2	H ₂ O	616.6	614.8	1.8	Pre-Test	<u>.021</u>	<u>10</u>	<u>CHF</u>			
		#3	Ø	515.8	515.3	0.5	Post-Test	<u>.001</u>	<u>10</u>	<u>CHF</u>			
		#4	SG	786.2	779.9	6.3	PRE-TEST CALIBRATION CHECK:						
		#5					Time	ΔH	Meter Reading	Meter Temp. In	Meter Temp. Out		
		Total					Init.						
		POST TEST INFO:							Final				
		Filter Appearance _____											
		Impingor Appearance _____											
Silica Gel Spent (Y/N) _____													

SAMPLE POINT	TIME	METER CONDITIONS			TEMPERATURES, °F							O ₂	VAC.	STATIC PRESS. iwg	CHAIN OF CUSTODY INFORMATION
		ΔP	ΔH	METER READING	STACK	PROBE	METER		OVEN	IMP. OUT					
							IN	OUT							
	1130		2.0	169.435			60	58		52		0		Impingers Loaded	
	1140		2.0	177.70			73	60		54		0		Impingers Recovered	
	1150		2.0	184.52			79	65		56		0		Filter Loaded	
	1205		2.2	194.605			80	69		56		0		Filter Recovered	
														Probe Wash	
TEST SUMMARY															
Calculated by: <u>CHF</u>															
Checked by: <u>CHF</u> ✓															
Sample Vol., c.f. <u>25,170</u> ✓															
Stack Press., iwg _____															
ΔH, iwg <u>2.0</u>															
ΔP, iwg _____															
Motor Temp., °F <u>68.0</u> ✓															
Stack Temp., °F _____															
Water Collected, g <u>23.6</u> ✓															
O ₂ /CO ₂ _____															
Comments: _____															

Appendix C.5
Total Hydrocarbon Lab Results



AtmAA Inc.

21354 Nordhoff St., Suite 113, Chatsworth, CA 91311 (818) 718-6070 • FAX (818) 718-9779

environmental consultants
laboratory services

January 6, 1993

LTR/005/93

Craig Fry
CARNOT
15991 Red Hill Ave. Suite 110
Tustin, CA 92680

JAN 11 1993

CARNOT

re: 54700

Dear Craig:

Please find enclosed the laboratory analysis report, quality assurance summary, and the original chain of custody form for four Tedlar bag samples received on December 31, 1992.

The samples were analyzed for methane and total gaseous non-methane organics.

Sincerely,

AtmAA, Inc.

Michael L. Porter
Laboratory Director

Encl.
MLP/krp



AtmAA Inc.

21354 Nordhoff St., Suite 113, Chatsworth, CA 91311 (818) 718-6070 • FAX (818) 718-9779

environmental consultants
laboratory services

LABORATORY ANALYSIS REPORT

Methane & Total Gaseous Non-Methane Organics
Analysis in Tedlar Bag Samples

Report Date: January 6, 1993
Client: CARNOT
P.O. No.: 9307
Project No.: 54700
Date Received: December 31, 1992
Date Analyzed: December 31, 1992

ANALYSIS DESCRIPTION

Methane and total gaseous non-methane organics were measured by flame ionization detection/total combustion analysis portion (FID/TCA).

AtmAA Lab No.:	93662-2	93662-3	93662-4	93662-5
Sample I.D.:	1-HC-In	1-HC-Out	2-HC-In	2-HC-Out
Components:	(Concentration in ppm, v/v)			
Methane	192000	<1	199000	<1
TGNMO	7040	15.8	7360	16.3

TGNMO is total gaseous non-methane organics measured and reported as ppm methane.

Michael L. Porter
Laboratory Director

QUALITY ASSURANCE SUMMARY
(Repeat Analysis)

Project No.: 54700
Date Received: December 31, 1992

Component:	Sample ID	Repeat Analysis		Mean Conc.	% Diff. From Mean
		Run #1	Run #2		
(Concentration in ppm, v/v)					
Methane	1-HC-In	192000	193000	192000	0.26
	1-HC-Out	<1	<1	---	---
	2-HC-Out	<1	<1	---	---
TGNMO	1-HC-In	7150	6930	7040	1.6
	1-HC-Out	15.6	16.1	15.8	1.6

A set of 4 Tedlar bag samples laboratory numbers, 93662-(2-5) was analyzed for methane and TGNMO. Agreement between repeat analyses is a measure of precision and is shown above in the column "% Difference from Mean." Repeat analyses are an important part of AtmAA's quality assurance program. The average % Difference from Mean for 3 repeat measurements from the sample set of 4 Tedlar bag samples is 1.6%.



**CARNOT
TEDLAR BAG SAMPLE DATA**

CLIENT/PROJECT: PVERG

DATE: 12-30-97 UNIT: 2 BY: RM

	BAG ID <u>797</u>	BAG ID <u>796</u>	BAG ID _____
TEST NO.	<u>1-THC-Out</u>	<u>2-THC-Out</u>	<u>1-HC-Out</u>
SAMPLE RATE			
START TIME	<u>1005</u>	<u>1122</u>	<u>0940</u>
STOP TIME	<u>1025</u>	<u>1142</u>	<u>1000</u>
SAMPLE LOCATION	<u>Stack</u>	<u>Stack</u>	<u>Stack</u>

DATE: _____ UNIT: _____ BY: _____

	BAG ID _____	BAG ID _____	BAG ID _____
TEST NO.	<u>2-HC-Out</u>		
SAMPLE RATE			
START TIME	<u>1145</u>		
STOP TIME	<u>1205</u>		
SAMPLE LOCATION	<u>Stack</u>		

NOTES:

**CARNOT
TEDLAR BAG SAMPLE DATA**

CLIENT/PROJECT: PVIER 6

DATE: 12-30-92 UNIT: PVIER 6 #1 BY: CHR

	BAG ID _____	BAG ID _____	BAG ID _____
TEST NO.	1-HC-In	2-HC-In	
SAMPLE RATE	Slpm	Slpm	
START TIME	950	1145	
STOP TIME	1010	1205	
SAMPLE LOCATION	Inlet	Inlet	

DATE: _____ UNIT: _____ BY: _____

	BAG ID _____	BAG ID _____	BAG ID _____
TEST NO.			
SAMPLE RATE			
START TIME			
STOP TIME			
SAMPLE LOCATION			

NOTES:

Appendix C.6
Trace Hydrocarbon and Trace Sulfur Results

TRACE ORGANIC SPECIES DESTRUCTION EFFICIENCY RESULTS
PVERG UNIT 1, TEST No. 1
DECEMBER 30, 1992

Flow rate, Inlet dscfm		3589		
Flow rate, Exhaust dscfm		14576		
Species	Inlet	Exhaust		Destruction Efficiency, %
Methylene Chloride:	200 ppb	ND< 0.2 ppb		
lb/hr	9.64E-03	< 3.91E-05		99.59
Chloroform:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.35E-02	< 5.50E-05		99.59
1,1,1 Trichloroethane:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.51E-02	< 6.15E-05		99.59
Carbon Tetrachloride:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.75E-02	< 7.09E-05		99.59
1,1-Dichloroethene:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.10E-02	< 4.47E-05		99.59
Trichloroethylene:	500 ppb	ND< 0.2 ppb		
lb/hr	3.73E-02	< 6.06E-05		99.84
Tetrachloroethylene:	400 ppb	ND< 0.2 ppb		
lb/hr	3.76E-02	< 7.64E-05		99.80
Chlorobenzene:	4700 ppb	ND< 0.4 ppb		
lb/hr	3.00E-01	< 1.04E-04		99.97
Vinyl Chloride:	1400 ppb	ND< 0.4 ppb		
lb/hr	4.97E-02	< 5.76E-05		99.88
m-Dichlorobenzene:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.67E-02	< 6.78E-05		99.59
o+p Dichlorobenzene:	1300 ppb	ND< 0.2 ppb		
lb/hr	1.08E-01	< 6.78E-05		99.94
1,1-Dichloroethane:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.12E-02	< 4.56E-05		99.59
1,2-Dichloroethane:	ND< 200 ppb	ND< 0.2 ppb		
lb/hr	< 1.12E-02	< 4.56E-05		99.59
Benzene:	11000 ppb	0.2 ppb		
lb/hr	4.88E-01	3.60E-05		99.99
Toluene:	24000 ppb	ND< 1 ppb		
lb/hr	1.25E+00	< 2.12E-04		99.98
Acetonitrile:	ND< 2000 ppb	ND< 2 ppb		
lb/hr	< 4.66E-02	< 1.89E-04		99.59
m-Xylenes:	25000 ppb	ND< 0.2 ppb		
lb/hr	1.51E+00	< 4.89E-05		100.00
o+p Xylenes:	21000 ppb	ND< 0.2 ppb		
lb/hr	1.27E+00	< 4.89E-05		100.00
1,2-Dibromoethane:	ND< 7000 ppb	ND< 7 ppb		
lb/hr	< 7.46E-01	< 3.03E-03		99.59
Benzyl Chloride:	ND< 5000 ppb	ND< 5 ppb		
lb/hr	< 3.59E-01	< 1.46E-03		99.59

ND< - indicates that the species was not detected. Values indicate the detection limit for this species and the concentration is less than the presented value.

NA - indicates that the destruction efficiency cannot be calculated because the concentration is below the detection limits.

TRACE ORGANIC SPECIES DESTRUCTION EFFICIENCY RESULTS
PVERG UNIT 1, TEST No. 2
DECEMBER 30, 1992

Flow rate, Inlet dscfm		3557	
Flow rate, Exhaust dscfm		14654	
Species	Inlet	Exhaust	Destruction Efficiency, %
Methylene Chloride:	ND< 200 ppb	0.4 ppb	
lb/hr	< 9.55E-03	7.87E-05	99.18
Chloroform:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.34E-02	< 5.53E-05	99.59
1,1,1 Trichloroethane:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.50E-02	< 6.18E-05	99.59
Carbon Tetrachloride:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.73E-02	< 7.13E-05	99.59
1,1-Dichloroethene:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.09E-02	< 4.49E-05	99.59
Trichloroethylene:	400 ppb	ND< 0.2 ppb	
lb/hr	2.96E-02	< 6.09E-05	99.79
Tetrachloroethylene:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.87E-02	< 7.68E-05	99.59
Chlorobenzene:	2400 ppb	ND< 0.4 ppb	
lb/hr	1.52E-01	< 1.04E-04	99.93
Vinyl Chloride:	1400 ppb	ND< 0.4 ppb	
lb/hr	4.92E-02	< 5.79E-05	99.88
m-Dichlorobenzene:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.65E-02	< 6.81E-05	99.59
o+p Dichlorobenzene:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.65E-02	< 6.81E-05	99.59
1,1-Dichloroethane:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.11E-02	< 4.59E-05	99.59
1,2-Dichloroethane:	ND< 200 ppb	ND< 0.2 ppb	
lb/hr	< 1.11E-02	< 4.59E-05	99.59
Benzene:	11000 ppb	0.1 ppb	
lb/hr	4.83E-01	1.81E-05	100.00
Toluene:	22000 ppb	ND< 1 ppb	
lb/hr	1.14E+00	< 2.13E-04	99.98
Acetonitrile:	ND< 2000 ppb	ND< 2 ppb	
lb/hr	< 4.62E-02	< 1.90E-04	99.59
m-Xylenes:	22000 ppb	ND< 0.2 ppb	
lb/hr	1.31E+00	< 4.92E-05	100.00
o+p Xylenes:	19000 ppb	ND< 0.2 ppb	
lb/hr	1.13E+00	< 4.92E-05	100.00
1,2-Dibromoethane:	ND< 7000 ppb	ND< 7 ppb	
lb/hr	< 7.40E-01	< 3.05E-03	99.59
Benzyl Chloride:	ND< 5000 ppb ¹	ND< 5 ppb	
lb/hr	< 3.56E-01	< 1.47E-03	99.59

ND< - indicates that the species was not detected. Values indicate the detection limit for this species and the concentration is less than the presented value.

NA - indicates that the destruction efficiency cannot be calculated because the concentration is below the detection limits.

PALOS VERDES LANDFILL

PVERG BOILER

Test Date: 12/30/92

	RAW GAS TEST #1	RAW GAS TEST #2	FLARE EXHAUST TEST #1	FLARE EXHAUST TEST #2	AMBIENT AIR AT FLARE STATION
JWPCP LAB SAMPLE NO.	JW95191	JW95192	JW95193	JW95194	N/A
COMPOUND					
1. HYDROGEN SULFIDE, ppm	21	22			
2. METHYL MERCAPTAN, ppm	0.5	0.5			
3. ETHYL MERCAPTAN, ppm	< 0.5	< 0.5			
4. CARBONYL SULFIDE, ppm	< 0.4	< 0.4			
5. CARBON DISULFIDE, ppm	< 0.6	< 0.6			
6. DIMETHYL SULFIDE, ppm	< 0.6	< 0.6			
7. DIMETHYL DISULFIDE, ppm	< 0.5	< 0.5			
8. PERMANENT GASES, TOT %	99.6	99.9	97.8	99	
9. OXYGEN (O2) %	11.2	11.2	5.99	5.81	
10. ARGON (AR) %	0.62	0.62	0.8	0.81	
11. NITROGEN (N2) %	54.6	54.4	79	78.9	
12. METHANE (CH4) %	17.3	17.5	< 0.02	< 0.02	
13. CARBON DIOXIDE (CO2) %	15.8	16.1	12	12.5	
14. METHYLENE CHLORIDE, ppb	200	< 200	< 0.2	0.4	
15. CHLOROFORM, ppb	< 200	< 200	< 0.2	< 0.2	
16. 1,1,1-TRICHLOROETHANE, ppb	< 200	< 200	< 0.2	< 0.2	
17. CARBON TETRACHLORIDE, ppb	< 200	< 200	< 0.2	< 0.2	
18. 1,1-DICHLOROETHENE, ppb	< 200	< 200	< 0.2	< 0.2	
19. TRICHLOROETHYLENE, ppb	500	400	< 0.2	< 0.2	
20. TETRACHLOROETHYLENE, ppb	400	< 200	< 0.2	< 0.2	
21. CHLOROBENZENE, ppb	4700	2400	< 0.4	< 0.4	
22. VINYL CHLORIDE, ppb	1400	1400	< 0.4	< 0.4	
23. M-DICHLOROENZENE, ppb	< 200	< 200	< 0.2	< 0.2	
24. 1,1-DICHLOROETHANE, ppb	< 200	< 200	< 0.2	< 0.2	
25. 1,2-DICHLOROETHANE, ppb	< 200	< 200	< 0.2	< 0.2	
26. BENZENE, ppb	11000	11000	0.2	0.1	
27. TOLUENE, ppb	24000	22000	< 1	1	
28. ACETONITRILE, ppb	< 2000	< 2000	< 2	2	
29. M-XYLENE, ppb	25000	22000	< 0.2	< 0.2	
30. O+P XYLENE, ppb	21000	19000	< 0.2	< 0.2	
31. 1,2-DIBROMOETHANE, ppb	< 7000	< 7000	< 7	7	
32. O+P DICHLOROENZENE, ppb	1300	< 200	< 0.2	< 0.2	
33. BENZYL CHLORIDE, ppb	< 5000	< 5000	< 5	5	

f/n: pv9212

Appendix C.7
Unit Operation Data

30DEC92 WEDNESDAY

PVERG B-501 PERFORMANCE DATA

TREND LOG

	J1-626 MEGAWATTS	GEN MEGAWATTS	J1-209 PARASITIC LOAD	PLANT BLR 1 LFG FLOW	FI-397 NAT GAS FLOW	BLR 1 LFG SUPPLY BTU	FR-303 SUPPLY
	J1-210 MEGAWATTS	SCF	FI-335	BLR 1 LFG FLOW	AR-304	LFG SUPPLY BTU	FR-303
	J1-626	J1-210	J1-209	FI-335	FI-397	AR-304	FR-303
	MW	MW	KW	KSCFM	SCFM	BTU/CF	KSCFM
30DEC92 00:01	9.0625	8.1563	908.00	3.6875	209.50	187.00	4.2344
00:16	9.0938	8.1875	910.00	3.6875	209.50	187.00	4.2344
00:31	9.0938	8.1875	908.00	3.6875	209.50	187.50	4.2344
00:46	9.0625	8.1563	910.00	3.6875	209.00	186.50	4.2344
01:01	9.0938	8.1875	910.00	3.6875	208.50	187.00	4.2344
01:16	9.0625	8.1563	908.00	3.6875	209.00	187.00	4.2344
01:31	9.0625	8.1563	912.00	3.6875	208.00	186.00	4.2344
01:46	9.0000	8.0938	906.00	3.6875	207.00	185.50	4.2344
02:01	9.0313	8.0938	908.00	3.6875	207.50	185.00	4.2344
02:16	9.0000	8.0938	902.00	3.6797	207.00	185.50	4.2188
02:31	9.0000	8.0938	904.00	3.6954	207.50	185.00	4.2344
02:46	9.0000	8.0938	906.00	3.6954	207.50	184.50	4.2344
03:01	8.9375	8.0313	904.00	3.6875	208.00	183.00	4.2344
03:16	8.7813	7.9663	908.00	3.6875	207.50	180.50	4.2344
03:31	8.5938	7.7188	896.00	3.6016	208.00	178.50	4.1563
03:46	8.5938	7.7188	892.00	3.6094	207.50	178.00	4.1719
04:01	8.5938	7.7188	896.00	3.6329	207.00	178.00	4.1875
04:16	8.5625	7.7032	892.00	3.6250	207.00	177.50	4.1719
04:31	8.5938	7.7032	896.00	3.6329	207.00	177.50	4.1875
04:46	8.5313	7.6407	890.00	3.6172	207.00	176.50	4.1719
05:01	8.5313	7.6407	892.00	3.6329	207.00	176.00	4.1719
05:16	8.5000	7.5938	892.00	3.6094	206.50	174.50	4.1719
05:31	8.4688	7.5938	894.00	3.6329	205.50	175.00	4.1719
05:46	8.5313	7.6250	891.00	3.6329	205.00	176.50	4.1719
06:01	8.7188	7.8438	878.00	3.6329	204.50	180.50	4.1563
06:16	8.8125	7.9844	864.00	3.6185	204.00	183.50	4.1719
06:31	8.9063	8.0313	872.00	3.6407	203.50	184.50	4.1875
06:46	8.9688	8.0625	878.00	3.6407	203.50	184.50	4.2188
07:01	9.0000	8.1250	892.00	3.6954	203.50	184.00	4.2657
07:16	9.0000	8.0938	884.00	3.6954	203.50	184.50	4.2657
07:31	8.9688	8.0625	890.00	3.6954	203.50	184.00	4.2500
07:46	8.9688	8.0625	890.00	3.7032	204.00	184.00	4.2657
08:01	8.9688	8.0625	886.00	3.6954	204.50	185.50	4.2657
08:16	8.9688	8.0625	886.00	3.6797	205.00	213.00	4.2500
08:31	9.0313	8.0938	908.00	3.6797	205.00	271.00	4.2500
08:46	9.0938	8.1563	910.00	3.7032	205.00	186.00	4.2813
09:01	9.1250	8.2188	914.00	3.7266	208.50	187.00	4.2969
09:16	9.1250	8.2188	922.00	3.7344	212.50	187.50	4.3125
09:31	9.1250	8.2188	922.00	3.7266	212.50	187.00	4.3125
09:46	9.0938	8.1875	918.00	3.7266	213.00	187.50	4.3125
10:01	9.1250	8.1875	926.00	3.7266	213.00	189.00	4.3125
10:16	9.1250	8.1875	922.00	3.7266	213.50	189.50	4.3125
10:31	9.1563	8.1875	960.00	3.7266	214.50	191.00	4.3282
10:46	9.1563	8.1875	960.00	3.7422	210.50	192.00	4.3125
11:01	9.1563	8.1875	962.00	3.7500	208.00	193.00	4.3282
11:16	9.1563	8.2188	962.00	3.7500	208.50	194.00	4.3282
11:31	9.2500	8.2813	958.00	3.7500	208.50	195.00	4.3282
11:46	9.2500	8.3438	926.00	3.7500	209.00	196.50	4.3438
12:01	9.2188	8.3125	914.00	3.6719	209.00	197.50	4.2969
12:16	9.2500	8.3438	916.00	3.6875	208.50	199.00	4.2969
12:31	9.2813	8.3750	912.00	3.7110	203.00	199.50	4.2969
12:46	9.3125	8.4063	912.00	3.7266	200.00	201.00	4.2969
13:01	9.3438	8.4375	908.00	3.7266	200.00	201.00	4.2969

Test 1

Test 2

30DEC92 WEDNESDAY

PVERG 8-501 DAILY EMISSIONS

TREND LOG

		RTU-501 B-501		NOX1-PPH B-501		A1-421 BLR 1		NOXC-1 B-	
		MBTU INPUT		TOTAL NOX-PPH		EXCESS O2		CORR NOX FDI	
		FI-523	BLR 1	A1-425	BLR 1	A1-426	BLR 1		
		STEAM FLOW		STACK NOX		STACK OXYGEN			
		RTU-501	FI-523	NOX1-PPH	A1-425	A1-421	A1-426	NOXC-1	
30DEC92		MMBTU	KLB/HR	PPH	PPM	%	%	PPM	
00:02		54.750	44.500	1.5743	22.063	1.7188	5.1407	25.125	
00:17		54.750	44.625	1.5743	22.063	1.7266	5.2188	25.188	
00:32		54.750	44.500	1.5625	21.875	1.7305	5.2344	25.000	
00:47		54.500	44.375	1.5704	22.125	1.7266	5.1407	25.125	
01:02		54.625	44.500	1.5586	21.813	1.7227	5.2344	24.938	
01:17		54.625	44.500	1.5743	22.125	1.7305	5.2188	25.250	
01:32		54.375	44.375	1.5625	21.938	1.7266	5.1563	25.000	
01:47		54.125	44.125	1.5430	21.875	1.7461	5.2500	25.000	
02:02		54.125	44.125	1.5547	21.938	1.7266	5.1719	25.063	
02:17		54.125	44.000	1.5586	22.063	1.7500	5.1563	25.125	
02:32		54.125	44.125	1.5625	22.125	1.7188	5.0938	25.063	
02:47		54.125	44.000	1.5586	22.063	1.7461	5.1563	25.063	
03:02		53.750	43.750	1.5274	21.813	1.7539	5.2657	25.000	
03:17		53.000	43.125	1.5235	22.063	1.7618	5.1250	25.063	
03:32		51.875	41.875	1.4688	21.875	1.7696	5.1719	24.938	
03:47		51.750	41.875	1.4766	22.000	1.7657	5.1719	25.063	
04:02		51.875	41.875	1.4649	21.688	1.7618	5.2657	24.938	
04:17		51.750	41.750	1.4766	21.938	1.7774	5.1563	25.000	
04:32		51.750	41.750	1.4649	21.813	1.7696	5.1094	24.688	
04:47		51.375	41.250	1.4571	21.813	1.7657	5.1563	24.813	
05:02		51.250	41.375	1.4610	21.813	1.7813	5.1250	24.813	
05:17		50.750	41.000	1.4180	21.375	1.7696	5.4844	24.875	
05:32		51.000	41.125	1.4414	21.625	1.7735	5.2813	24.875	
05:47		51.375	41.250	1.4727	22.000	1.7813	5.1250	25.000	
06:02		52.500	42.375	1.4961	21.938	1.7383	5.1407	24.938	
06:17		53.000	43.125	1.5196	22.000	1.7383	5.2500	25.125	
06:32		53.125	43.250	1.5235	21.938	1.7539	5.2188	25.125	
06:47		53.125	43.375	1.5430	22.250	1.7383	5.1407	25.313	
07:02		53.875	44.000	1.5743	22.313	1.7422	5.0625	25.250	
07:17		53.875	44.000	1.5743	22.375	1.7461	5.2500	25.563	
07:32		53.750	43.875	1.5586	22.188	1.7422	5.2813	25.438	
07:47		53.875	43.875	1.5821	22.438	1.7266	5.1250	25.500	
08:02		51.250	43.750	1.3633	19.375	1.7305	7.2188	22.438	
08:17		66.750	43.625	2.2579	32.125	1.7266	9.8438	27.938	
08:32		68.750	43.750	2.7032	38.375	1.7579	7.1094	33.625	
08:47		54.375	44.000	4.2032	59.250	1.7461	0.5313	51.000	
09:02		55.000	44.500	2.0235	26.438	1.7110	4.2344	29.438	
09:17		55.375	44.500	1.5664	21.813	1.7188	5.2032	24.875	
09:32		55.250	44.375	1.5391	21.438	1.7539	5.2032	24.500	
09:47		55.375	44.250	1.5508	21.625	1.7110	5.0938	24.563	
10:02		55.875	44.375	1.5313	21.250	1.7500	5.2032	24.250	
10:17		56.000	44.375	1.5274	21.250	1.7071	5.0782	24.063	
10:32		56.250	44.500	1.5235	21.125	1.7383	5.1250	24.000	
10:47		56.500	44.500	1.5118	20.875	1.7305	5.2344	23.875	
11:02		56.625	44.625	1.4922	20.500	1.7110	5.2188	23.438	
11:17		56.750	44.875	1.5000	20.563	1.7149	5.1563	23.438	
11:32		57.250	45.125	1.4883	20.250	1.7188	5.2188	23.125	
11:47		57.375	45.125	1.5079	20.500	1.7383	5.2032	23.375	
12:02		56.750	44.375	1.4805	20.500	1.7149	5.1719	23.375	
12:17		57.125	44.875	1.5000	20.625	1.7071	5.1563	23.500	
12:32		57.250	45.000	1.4922	20.313	1.7110	5.1719	23.188	
12:47		57.625	45.125	1.5274	20.813	1.7188	5.2344	23.813	
13:02		57.750	45.375	1.5274	20.688	1.7227	5.2032	23.625	

APPENDIX D
CALCULATIONS

Appendix D.1
General Emissions Calculations

EMISSION CALCULATIONS

1. Sample Volume and Isokinetics

- a. Sample gas volume, dscf

$$V_{m\ std} = 0.03342 V_m \left(P_{bar} + \frac{H}{13.6} \right) \left(\frac{T_{ref}}{T_m} \right) (Y)$$

- b. Water vapor volume, scf

$$V_{w\ std} = 0.0472 V_{lc} \left(\frac{T_{ref}}{528 \text{ } ^\circ R} \right)$$

- c. Moisture content, nondimensional

$$B_{wo} = \frac{V_{w\ std}}{V_{m\ std} + V_{w\ std}}$$

- d. Stack gas molecular weight, lb/lb mole

$$MW_{dry} = 0.44 (\%CO_2) + 0.32 (\%O_2) + 0.28 (\%N_2)$$

$$MW_{wet} = MW_{dry} (1 - B_{wo}) + 18 (B_{wo})$$

- e. Absolute stack pressure, iwg

$$P_s = P_{bar} + \frac{P_{sg}}{13.6}$$

- f. Stack velocity, ft/sec

$$V_s = 2.90 C_p \sqrt{\Delta P T_s} \sqrt{\left(\frac{29.92}{P_s} \right) \left(\frac{28.95}{MW_{wet}} \right)}$$

- g. Actual stack flow rate, wacfm

$$Q = (V_s)(A_s)(60)$$

- h. Standard stack gas flow rate, dscfm

$$Q_{std} = Q (1 - B_{wo}) \left(\frac{T_{ref}}{T_s} \right) \left(\frac{P_s}{29.92} \right)$$

- i. Percent isokinetic

$$I = \left(\frac{17.32 (T_p) (V_{m\ std})}{(1 - B_{wo}) (\Theta) (V_p) (P_p) (D_n^2)} \right) \left(\frac{528 \text{ } ^\circ R}{T_{ref}} \right)$$

2. Particulate Emissions

- a. Grain loading, gr/dscf

$$C = 0.01543 \left(\frac{M_n}{V_{m \text{ std}}} \right)$$

- b. Grain loading at 12% CO
- ₂
- , gr/dscf

$$C_{12\% \text{ CO}_2} = C \left(\frac{12}{\% \text{ CO}_2} \right)$$

- c. Mass emissions, lb/hr

$$M = C(Q_{\text{std}}) \frac{(60 \text{ min/hr})}{(7000 \text{ gr/lb})}$$

3. Gaseous Emissions, lb/hr

$$M = (\text{ppm}) (10^{-6}) \left(\frac{MW_i \text{ lb/lb mole}}{SV} \right) (Q_{\text{std}}) (60 \text{ min/hr})$$

where,

SV = specific molar volume of an ideal gas:

$$SV = 385.3 \text{ ft}^3/\text{lb mole} \quad \text{for } T_{\text{ref}} = 528 \text{ }^\circ\text{R}$$

$$SV = 379.5 \text{ ft}^3/\text{lb mole} \quad \text{for } T_{\text{ref}} = 520 \text{ }^\circ\text{R}$$

4. Emissions Rates, lb/10⁶ Btu

- a. Fuel factor at 68 °F, dscf/10
- ⁶
- Btu at 0% O
- ₂

$$F_{68} = \frac{10^6 [3.64 (\%H) + 1.53 (\%C) + 0.14 (\%N) + 0.57 (\%S) - 0.46 (\%O_2 \text{ fuel})]}{HHV, \text{ Btu/lb}}$$

- b. Fuel factor at 60 °F

$$F_{60} = F_{68} \left(\frac{520 \text{ }^\circ\text{R}}{528 \text{ }^\circ\text{R}} \right)$$

- c. Gaseous Emissions factor

$$\left(\frac{\text{lb}}{10^6 \text{ Btu}} \right)_i = (\text{ppm})_i (10^{-6}) \left(\frac{MW_i \text{ lb}}{\text{lb mole}} \right) \left(\frac{1}{SV} \right) (F) \left(\frac{20.9}{20.9 - \%O_2} \right)$$

- d. Particulate emission factor

$$\left(\frac{\text{lb}}{10^6 \text{ Btu}} \right) = C \left(\frac{1 \text{ lb}}{7000 \text{ gr}} \right) (F) \left(\frac{20.9}{20.9 - \%O_2} \right)$$

Nomenclature:

A_s	=	stack area, ft ²
B_{wo}	=	flue gas moisture content
$C_{12\% CO_2}$	=	particulate grain loading, gr/dscf corrected to 12% CO ₂
C	=	particulate grain loading, gr/dscf
C_p	=	pitot calibration factor, dimensionless
D_n	=	nozzle diameter, in.
F	=	fuel F factor, dscf/10 ⁶ Btu at 0% O ₂
H	=	orifice pressure differential, iwg
I	=	% isokinetics
M_n	=	mass of collected particulate, mg
M_i	=	mass emissions of species i, lb/hr
MW	=	molecular weight of flue gas
MW_i	=	molecular weight of species i:
		NO _x : 46
		CO : 28
		SO _x : 64
		HC : 16
θ	=	sample time, min.
ΔP	=	average velocity head, iwg = $(\sqrt{\Delta P})^2$
P_{bar}	=	barometric pressure, in.Hg
P_s	=	stack absolute pressure, in.Hg
P_{sg}	=	stack static pressure, iwg

Nomenclature (Continued):

Q	=	wet stack gas flow rate at actual conditions, wacfm
Q_{sd}	=	dry stack gas flow rate at standard conditions, dscfm
SV	=	specific molar volume of an ideal gas at standard conditions, ft ³ /lb mole
T_m	=	meter temperature, °R
T_{ref}	=	reference temperature, °R
T_s	=	stack temperature, °R
V_s	=	stack velocity, ft/sec
V_{lc}	=	volume of liquid collected in impingers, ml
V_m	=	dry meter volume uncorrected, dcf
$V_{m\ sd}$	=	dry meter volume at standard conditions, dscf
$V_{w\ sd}$	=	volume of water vapor at standard conditions, scf
Y	=	meter calibration coefficient

Appendix D.2
Sample Calculations

CARNOT

SUBJECT Sample Calculations for Flow Rate, JOB NO. _____
Moisture and Particulate Matter SHEET NO. 1 of 5
COMPUTED BY C. Fry DATE 1-18-93 CHECKED BY _____ DATE _____

Test No. 2-PM-PV

See Appendix D.1 for
Nomenclature

Note: $\text{avg } \Delta P = (\sum \sqrt{\Delta P})^2$

$V_{\text{std}} = \text{Standard Sample Volume Metered (scf)}$

$$V_{\text{m std}} = 0.03342 (V_{\text{m}}) \left(P_{\text{Bar}} + \frac{\Delta H}{13.6} \right) \left(\frac{T_{\text{ref}}}{T_{\text{m}}} \right) Y$$

$$V_{\text{m std}} = 0.03342 (33.677) \left(29.96 + \frac{0.955}{13.6} \right) \frac{520}{(79.78 + 460)}$$

$$V_{\text{m std}} = 32.526$$

$V_{\text{w std}} = \text{Water Vapor Volume (scf)}$

$$V_{\text{w std}} = 0.0472 (V_{\text{w}}) \left(\frac{T_{\text{ref}}}{528^{\circ}\text{R}} \right)$$

$$V_{\text{w std}} = 0.0472 (120.0) \frac{520}{528}$$

$$V_{\text{w std}} = 5.57818$$

CARNOT

SUBJECT Sample Calculations Continued

JOB NO. _____

SHEET NO. 2 of 5

COMPUTED BY C. Fry

DATE _____

CHECKED BY _____

DATE _____

$B_{wo} = \text{Moisture Content}$

$$B_{wo} = \frac{V_{w \text{ std}}}{V_{m \text{ std}} + V_{w \text{ std}}} \times 100$$

$$B_{wo} = \frac{5.57818}{32.526 + 5.57818}$$

$$B_{wo} = 0.146 \quad \text{or} \quad \times 100 = 14.6 \%$$

$MW = \text{Stack Gas Molecular Weight}$

$$MW_{\text{dry}} = 0.44(\% \text{CO}_2) + 0.32(\% \text{O}_2) + 0.28(\% \text{N}_2)$$

$$MW_{\text{dry}} = 0.44(12.83) + 0.32(5.39) + 0.28(81.8)$$

$$MW_{\text{dry}} = 30.274$$

$$MW_{\text{wet}} = MW_{\text{dry}}(1 - B_{wo}) + 18(B_{wo})$$

$$MW_{\text{wet}} = 30.274(1 - 0.146) + 18(0.146)$$

$$MW_{\text{wet}} = 28.482$$

CARNOT

SUBJECT Sample Calculations, Continued JOB NO. _____

SHEET NO. 3 of 5

COMPUTED BY C. Fry DATE _____ CHECKED BY _____ DATE _____

$P_s = \text{Absolute Stack Pressure}$

$$P_s = P_{\text{bar}} + \frac{P_{\text{st}}}{13.6}$$

$$P_s = 29.96 + \frac{.1181}{13.6}$$

$$P_s = 29.969$$

$V_s = \text{Stack Velocity ft/sec}$

$$V_s = 2.90 (C_p) \sqrt{(\Delta P)(T_s)} \sqrt{\left(\frac{29.92}{P_s}\right) \left(\frac{28.95}{MW_{\text{wet}}}\right)}$$

$$V_s = 2.90 (.84) \sqrt{(4750)(536.75 + 460)} \sqrt{\frac{29.92}{29.969} \frac{28.95}{28.482}}$$

$$V_s = 2.90 (.84) (19.45395) / 1.00736$$

$$V_s = 47.738$$

Note: $T_s = \text{Temp Stack } F^{\circ} + 460$

139696

CARNOT

SUBJECT Sample Calculations, Continued

JOB NO. _____

SHEET NO. 4 of 5

COMPUTED BY _____

DATE _____

CHECKED BY _____

DATE _____

$Q = \text{Stack Flow Rate, Wet actual (wacfm)}$

$$Q = (V_s)(A_s) 60 \text{ sec/min}$$

$A_s = \text{Stack Area}$

$$Q = (47.738)(9.168) 60$$

$$Q = 26259.9$$

$Q_{std} = \text{Standard Stack Flow Rate, dscfm}$

$$Q_{std} = Q (1 - B_{wo}) \left(\frac{T_{ref}}{T_s} \right) \left(\frac{P_s}{29.92} \right)$$

$$Q_{std} = 26,259.9 (1 - .146) \frac{520}{460 + 336.75} \frac{29.969}{29.92}$$

$$Q_{std} = 14660$$

CARNOT

SUBJECT Sample Calculations, Continued JOB NO. _____

SHEET NO. 5 of 5

COMPUTED BY C. Fry DATE _____ CHECKED BY _____ DATE _____

$C = \text{particulate grain loading } \text{gr/dscf}$

$$C = 0.01543 \left(\frac{M_n}{V_{mstd}} \right)$$

$$C = 0.01543 \left(\frac{4.30}{32526} \right)$$

$$C = 0.0020$$

$M = \text{Mass Emissions } \text{lb/hr}$

$$M = (C \times Q_{std}) \frac{60 \text{ min/hr}}{7000 \text{ gr/lb}}$$

$$M = 0.0020 \frac{14660 \times 60}{7000}$$

$$M = 0.256$$

Appendix C.2
CEM Data and Strip Charts